

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110718\
 Data File : VN052201.D
 Acq On : 7 Nov 2018 11:50
 Operator : MD\SY
 Sample : J5818-02 5X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 EFF-WW

Quant Time: Nov 07 12:12:47 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N110118W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 02 05:39:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.20	128	65698	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.59	114	370720	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	311919	30.00	ug/l	0.00

System Monitoring Compounds

27) 1,2-Dichloroethane-d4	8.03	65	140200	28.75	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	95.83%
60) 4-Bromofluorobenzene	12.40	95	129667	25.99	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	86.63%
63) Toluene-d8	10.09	98	444269	30.29	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.97%

Target Compounds

					Ovalue
15) Acetone	3.82	58	367198	1145.649	ug/l 89
20) Diisopropyl ether	5.95	45	691803	51.609	ug/l 90
30) 2-Butanone	6.84	43	11038	7.042	ug/l # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110718\
Data File : VN052201.D
Acq On : 7 Nov 2018 11:50
Operator : MD\SY
Sample : J5818-02 5X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
EFF-WW

Quant Time: Nov 07 12:12:47 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N110118W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 02 05:39:38 2018
Response via : Initial Calibration

