

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110718\
 Data File : VN052208.D
 Acq On : 7 Nov 2018 15:38
 Operator : MD\SY
 Sample : J5844-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 14 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 MH-A-B

Quant Time: Nov 08 09:19:27 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 05 22:32:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	428903	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	710211	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	593637	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	196926	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	251658	49.48	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.96%	
35) Dibromofluoromethane	7.59	113	222872	50.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.08%	
50) Toluene-d8	10.09	98	854593	50.13	ug/l	0.00
Spiked Amount	50.000		Recovery	=	100.26%	
62) 4-Bromofluorobenzene	12.40	95	247766	44.27	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.54%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110718\
Data File : VN052208.D
Acq On : 7 Nov 2018 15:38
Operator : MD\SY
Sample : J5844-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 14 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
MH-A-B

Quant Time: Nov 08 09:19:27 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110518W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 05 22:32:25 2018
Response via : Initial Calibration

