

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110719\
 Data File : VN059117.D
 Acq On : 7 Nov 2019 13:20
 Operator : JC/SP
 Sample : K5704-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WGGM-19-WW-TSF-039

Quant Time: Nov 08 02:28:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N101119W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Oct 11 04:26:47 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.17	128	72457	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.57	114	387620	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.40	117	334297	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.00	65	175377	28.52	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	95.07%
60) 4-Bromofluorobenzene	12.40	95	141729	24.24	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	80.80%
63) Toluene-d8	10.08	98	472197	29.92	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	99.73%

Target Compounds					Ovalue
91) Naphthalene	15.13	128	9541	0.606 ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110719\
Data File : VN059117.D
Acq On : 7 Nov 2019 13:20
Operator : JC/SP
Sample : K5704-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WGGM-19-WW-TSF-039

Quant Time: Nov 08 02:28:58 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N101119W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Oct 11 04:26:47 2019
Response via : Initial Calibration

