

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110719\
 Data File : VN059121.D
 Acq On : 7 Nov 2019 14:48
 Operator : JC/SP
 Sample : K5675-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TR-04-110619

Quant Time: Nov 08 02:56:43 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 07 04:00:41 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	483882	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	741346	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	660540	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	242745	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	293229	49.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.60%	
35) Dibromofluoromethane	7.57	113	227414	51.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.74%	
50) Toluene-d8	10.08	98	911313	52.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.54%	
62) 4-Bromofluorobenzene	12.40	95	290940	46.28	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.56%	
Target Compounds						
20) Methylene Chloride	4.53	84	60313	11.192	ug/l	97
43) Isopropyl Acetate	8.14	43	122618	9.977	ug/l #	89
95) Naphthalene	15.13	128	29966	5.282	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110719\
Data File : VN059121.D
Acq On : 7 Nov 2019 14:48
Operator : JC/SP
Sample : K5675-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TR-04-110619

Quant Time: Nov 08 02:56:43 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110619W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 07 04:00:41 2019
Response via : Initial Calibration

