

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110722\
 Data File : VN075185.D
 Acq On : 07 Nov 2022 13:54
 Operator : JC\MD
 Sample : VN1107WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1107WBL01

Quant Time: Nov 08 01:08:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	95730	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	183744	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	157613	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	61310	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	74984	58.134	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.260%
35) Dibromofluoromethane	8.171	113	60764	56.332	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	112.660%
50) Toluene-d8	10.571	98	214631	51.286	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.580%
62) 4-Bromofluorobenzene	12.853	95	67333	45.802	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.600%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110722\
Data File : VN075185.D
Acq On : 07 Nov 2022 13:54
Operator : JC\MD
Sample : VN1107WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1107WBL01

Quant Time: Nov 08 01:08:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

