

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110722\
 Data File : VN075190.D
 Acq On : 07 Nov 2022 16:09
 Operator : JC\MD
 Sample : PB148848ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148848ZHE#02

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/08/2022
 Supervised By : Mahesh Dadoda 11/08/2022

Quant Time: Nov 08 01:10:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	94564	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	191459	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	176009	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	69443	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	74606	58.554	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 117.100%			
35) Dibromofluoromethane	8.171	113	59392	52.842	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.680%			
50) Toluene-d8	10.571	98	230755	52.917	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 105.840%			
62) 4-Bromofluorobenzene	12.853	95	75732	49.439	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 98.880%			
Target Compounds						
16) Acetone	4.453	43	8733	16.187	ug/l	99
18) Methyl Acetate	5.036	43	3964	2.249	ug/l	98
20) Methylene Chloride	5.283	84	13535	11.252	ug/l	89
37) Ethyl Acetate	7.571	43	17534	9.188	ug/l	96
43) Isopropyl Acetate	8.694	43	143332m	48.422	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110722\
Data File : VN075190.D
Acq On : 07 Nov 2022 16:09
Operator : JC\MD
Sample : PB148848ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148848ZHE#02

Manual Integrations
APPROVED

Reviewed By : John Carlone 11/08/2022
Supervised By : Mahesh Dadoda 11/08/2022

Quant Time: Nov 08 01:10:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

