

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110722\
 Data File : VN075194.D
 Acq On : 07 Nov 2022 17:46
 Operator : JC\MD
 Sample : N5479-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TUL-ROS-1104

Quant Time: Nov 08 01:12:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/08/2022
 Supervised By : Mahesh Dadoda 11/08/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	99959	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	205127	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	186055	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	73932	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	79436	58.980	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 117.960%			
35) Dibromofluoromethane	8.177	113	62268	51.709	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.420%			
50) Toluene-d8	10.571	98	243796	52.183	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.360%			
62) 4-Bromofluorobenzene	12.853	95	81422	49.612	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 99.220%			
Target Compounds					Qvalue	
16) Acetone	4.447	43	9468	16.602	ug/l	96
18) Methyl Acetate	5.053	43	4117	2.210	ug/l	98
20) Methylene Chloride	5.283	84	4778	3.758	ug/l	85
37) Ethyl Acetate	7.571	43	17793	8.703	ug/l	98
43) Isopropyl Acetate	8.688	43	130342m	41.100	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110722\
Data File : VN075194.D
Acq On : 07 Nov 2022 17:46
Operator : JC\MD
Sample : N5479-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TUL-ROS-1104

Quant Time: Nov 08 01:12:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/08/2022
Supervised By :Mahesh Dadoda 11/08/2022

