

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110722\
 Data File : VN075198.D
 Acq On : 07 Nov 2022 19:23
 Operator : JC\MD
 Sample : N5491-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Quant Time: Nov 08 01:13:40 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/08/2022
 Supervised By : Mahesh Dadoda 11/08/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	85253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	177195	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	164652	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	62198	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	70203	61.116	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 122.240%			
35) Dibromofluoromethane	8.171	113	55348	53.208	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 106.420%			
50) Toluene-d8	10.571	98	211853	52.494	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.980%			
62) 4-Bromofluorobenzene	12.853	95	68995	48.667	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 97.340%			
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	11419	23.477	ug/l #	88
18) Methyl Acetate	5.036	43	5575	3.509	ug/l	94
20) Methylene Chloride	5.289	84	4932	4.548	ug/l #	79
37) Ethyl Acetate	7.571	43	17342	9.819	ug/l	98
43) Isopropyl Acetate	8.683	43	139360m	50.870	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110722\
Data File : VN075198.D
Acq On : 07 Nov 2022 19:23
Operator : JC\MD
Sample : N5491-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Quant Time: Nov 08 01:13:40 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/08/2022
Supervised By :Mahesh Dadoda 11/08/2022

