

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110723\
 Data File : VN079965.D
 Acq On : 07 Nov 2023 15:22
 Operator : JC\MD
 Sample : 05222-05DL 5X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FSND-MW-29-20231101DL

Quant Time: Nov 08 04:29:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	304993	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	606583	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	621200	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	237015	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	249927	53.786	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.580%
35) Dibromofluoromethane	8.171	113	208251	48.508	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.020%
50) Toluene-d8	10.571	98	742322	46.969	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	93.940%
62) 4-Bromofluorobenzene	12.853	95	259813	49.567	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.140%
Target Compounds						Qvalue
44) Trichloroethene	9.353	130	164703	36.152	ug/l	99
64) Tetrachloroethene	11.112	164	16893	4.199	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110723\
Data File : VN079965.D
Acq On : 07 Nov 2023 15:22
Operator : JC\MD
Sample : 05222-05DL 5X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FSND-MW-29-20231101DL

Quant Time: Nov 08 04:29:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

