

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110723\
 Data File : VN079974.D
 Acq On : 07 Nov 2023 18:58
 Operator : JC\MD
 Sample : 05107-30 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1077

Quant Time: Nov 08 04:33:45 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	282127	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	585034	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	614579	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	251702	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	246044	57.242	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.480%
35) Dibromofluoromethane	8.171	113	203324	49.105	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.200%
50) Toluene-d8	10.571	98	741152	48.622	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.240%
62) 4-Bromofluorobenzene	12.853	95	270382	53.484	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	106.960%
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	10270	6.802	ug/l #	81
40) Benzene	8.612	78	475791	26.869	ug/l	96
43) Isopropyl Acetate	8.706	43	14351	1.509	ug/l #	74
52) Toluene	10.635	92	170906	16.280	ug/l	100
68) m/p-Xylenes	12.071	106	27851	3.219	ug/l	90
69) o-Xylene	12.400	106	19140	2.297	ug/l	95
95) Naphthalene	15.647	128	203607	15.374	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110723\
Data File : VN079974.D
Acq On : 07 Nov 2023 18:58
Operator : JC\MD
Sample : 05107-30 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1077

Quant Time: Nov 08 04:33:45 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

