

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110723\
 Data File : VN079976.D
 Acq On : 07 Nov 2023 19:46
 Operator : JC\MD
 Sample : 05107-32 20X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 1079

Quant Time: Nov 08 04:34:53 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	283655	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	584093	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	616123	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	256733	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	243925	56.443	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 112.880%	
35) Dibromofluoromethane	8.165	113	204537	49.477	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 98.960%	
50) Toluene-d8	10.571	98	727228	47.785	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.580%	
62) 4-Bromofluorobenzene	12.853	95	271243	53.741	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 107.480%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	9867	6.500	ug/l #	71
40) Benzene	8.612	78	195383	11.052	ug/l	96
43) Isopropyl Acetate	8.694	43	11945	1.258	ug/l #	76
52) Toluene	10.635	92	102813	9.809	ug/l	96
68) m/p-Xylenes	12.071	106	20021	2.308	ug/l	94
69) o-Xylene	12.400	106	13757	1.647	ug/l	89
95) Naphthalene	15.647	128	210632	15.593	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110723\
Data File : VN079976.D
Acq On : 07 Nov 2023 19:46
Operator : JC\MD
Sample : 05107-32 20X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
1079

Quant Time: Nov 08 04:34:53 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

