

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110723\
 Data File : VN079981.D
 Acq On : 07 Nov 2023 21:45
 Operator : JC\MD
 Sample : 05102-11DL 200X
 Misc : 2.70g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 67S-SB16-1617DL

Manual Integrations
 APPROVED

Reviewed By :John Carlone 11/08/2023
 Supervised By :Mahesh Dadoda 11/08/2023

Quant Time: Nov 08 04:38:04 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	254900	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	516359	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	540076	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	222132	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	230458	59.343	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 118.680%			
35) Dibromofluoromethane	8.171	113	192420	52.652	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 105.300%			
50) Toluene-d8	10.571	98	680350	50.569	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.140%			
62) 4-Bromofluorobenzene	12.853	95	255512	57.264	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 114.520%			
Target Compounds						
27) cis-1,2-Dichloroethene	7.494	96	27939	7.633	ug/l	95
31) Cyclohexane	8.230	56	13741m	2.678	ug/l	
39) Methylcyclohexane	9.600	83	5288	1.077	ug/l #	80
40) Benzene	8.612	78	18796	1.203	ug/l	99
52) Toluene	10.635	92	1156198	124.784	ug/l	99
64) Tetrachloroethene	11.106	164	4128	1.180	ug/l #	82
67) Ethyl Benzene	11.965	91	351387	17.216	ug/l	92
68) m/p-Xylenes	12.070	106	539879	71.005	ug/l	91
69) o-Xylene	12.400	106	164722	22.498	ug/l	93
78) n-propylbenzene	13.041	91	23753	1.230	ug/l	95
84) 1,2,4-Trimethylbenzene	13.488	105	37180	2.881	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110723\
Data File : VN079981.D
Acq On : 07 Nov 2023 21:45
Operator : JC\MD
Sample : 05102-11DL 200X
Misc : 2.70g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
67S-SB16-1617DL

Manual Integrations
APPROVED

Reviewed By : John Carlone 11/08/2023
Supervised By : Mahesh Dadoda 11/08/2023

Quant Time: Nov 08 04:38:04 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

