

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110723\
 Data File : VN079959.D
 Acq On : 07 Nov 2023 12:33
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Nov 08 04:25:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John
 Carlone
 11/08/2023
 Supervised By :Mahesh
 Dadoda
 11/08/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	331041	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	567705	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	515166	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	243820	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	286413	56.788	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 113.580%			
35) Dibromofluoromethane	8.171	113	220503	54.879	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 109.760%			
50) Toluene-d8	10.571	98	821272	55.523	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 111.040%			
62) 4-Bromofluorobenzene	12.853	95	302200	61.602	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 123.200%#			
Target Compounds						Qvalue
2) Dichlorodifluoromethane	2.124	85	193075	48.420	ug/l	99
3) Chloromethane	2.359	50	225362	45.205	ug/l	99
4) Vinyl Chloride	2.512	62	281870	54.003	ug/l	100
5) Bromomethane	2.936	94	156967	49.495	ug/l	96
6) Chloroethane	3.112	64	173256	42.009	ug/l	100
7) Trichlorofluoromethane	3.495	101	338776	52.184	ug/l	99
8) Diethyl Ether	3.959	74	129233	52.458	ug/l	68
9) 1,1,2-Trichlorotrifluo...	4.371	101	191590	50.909	ug/l	99
10) Methyl Iodide	4.589	142	227226	52.872	ug/l	99
11) Tert butyl alcohol	5.524	59	179699	279.111	ug/l	98
12) 1,1-Dichloroethene	4.342	96	178907	50.208	ug/l #	82
13) Acrolein	4.183	56	142686	292.927	ug/l	94
14) Allyl chloride	5.024	41	296211	54.275	ug/l	90
15) Acrylonitrile	5.718	53	571249	284.715	ug/l	100
16) Acetone	4.424	43	529354	298.802	ug/l	93
17) Carbon Disulfide	4.706	76	463185	45.064	ug/l	99
18) Methyl Acetate	5.024	43	482763	56.958	ug/l #	84
19) Methyl tert-butyl Ether	5.794	73	692112	56.416	ug/l	95
20) Methylene Chloride	5.277	84	219543	49.113	ug/l #	83
21) trans-1,2-Dichloroethene	5.789	96	199995	50.325	ug/l #	78
22) Diisopropyl ether	6.671	45	730519	57.759	ug/l #	90
23) Vinyl Acetate	6.606	43	2057770	294.961	ug/l #	87
24) 1,1-Dichloroethane	6.571	63	418782	54.477	ug/l	96
25) 2-Butanone	7.483	43	792449	305.465	ug/l #	80
26) 2,2-Dichloropropane	7.494	77	360730	56.740	ug/l	99
27) cis-1,2-Dichloroethene	7.488	96	248261	52.224	ug/l	86
28) Bromochloromethane	7.818	49	182041	48.592	ug/l #	76
29) Tetrahydrofuran	7.835	42	506583	300.137	ug/l #	76
30) Chloroform	7.965	83	432428	51.787	ug/l	91
31) Cyclohexane	8.259	56	331135	49.686	ug/l	89
32) 1,1,1-Trichloroethane	8.171	97	374066	54.979	ug/l	95
36) 1,1-Dichloropropene	8.371	75	314768	54.866	ug/l	95
37) Ethyl Acetate	7.565	43	316262	60.108	ug/l #	92
38) Carbon Tetrachloride	8.365	117	319248	55.215	ug/l	97
39) Methylcyclohexane	9.606	83	301238	55.795	ug/l #	87
40) Benzene	8.612	78	929678	54.104	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110723\
 Data File : VN079959.D
 Acq On : 07 Nov 2023 12:33
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :

Quant Time: Nov 08 04:25:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :John

Carlone

11/08/2023

Supervised By :Mahesh

Dadoda

11/08/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)	
41) Methacrylonitrile	7.783	41	173356	62.046	ug/l	#	84
42) 1,2-Dichloroethane	8.671	62	347589	57.910	ug/l		97
43) Isopropyl Acetate	8.694	43	548586	59.454	ug/l	#	89
44) Trichloroethene	9.353	130	219415	51.459	ug/l		95
45) 1,2-Dichloropropane	9.624	63	253949	57.830	ug/l		97
46) Dibromomethane	9.712	93	164253	54.780	ug/l		95
47) Bromodichloromethane	9.894	83	345643	54.052	ug/l		99
48) Methyl methacrylate	9.682	41	258415	63.939	ug/l	#	79
49) 1,4-Dioxane	9.700	88	88855	1244.214	ug/l	#	82
51) 4-Methyl-2-Pentanone	10.447	43	1653737	318.345	ug/l		91
52) Toluene	10.635	92	575230	56.468	ug/l		99
53) t-1,3-Dichloropropene	10.841	75	380331	60.670	ug/l		97
54) cis-1,3-Dichloropropene	10.318	75	400939	58.678	ug/l	#	90
55) 1,1,2-Trichloroethane	11.018	97	231931	56.605	ug/l		96
56) Ethyl methacrylate	10.876	69	373782	63.955	ug/l	#	85
57) 1,3-Dichloropropane	11.165	76	415660	57.553	ug/l		99
58) 2-Chloroethyl Vinyl ether	10.165	63	860068	290.350	ug/l		90
59) 2-Hexanone	11.200	43	1219463	306.860	ug/l		90
60) Dibromochloromethane	11.365	129	248177	56.586	ug/l		100
61) 1,2-Dibromoethane	11.476	107	231024	56.177	ug/l		98
64) Tetrachloroethene	11.106	164	172207	51.613	ug/l		97
65) Chlorobenzene	11.894	112	608925	54.490	ug/l		99
66) 1,1,1,2-Tetrachloroethane	11.965	131	231045	57.385	ug/l		96
67) Ethyl Benzene	11.965	91	1123967	57.729	ug/l		95
68) m/p-Xylenes	12.076	106	837812	115.516	ug/l		92
69) o-Xylene	12.400	106	407565	58.357	ug/l		91
70) Styrene	12.412	104	687988	61.297	ug/l		97
71) Bromoform	12.582	173	172581	60.365	ug/l	#	99
73) Isopropylbenzene	12.700	105	1040007	56.203	ug/l		100
74) N-amyl acetate	12.500	43	461363	57.121	ug/l		90
75) 1,1,2,2-Tetrachloroethane	12.941	83	352338	53.703	ug/l		99
76) 1,2,3-Trichloropropane	12.994	75	285121m	49.810	ug/l		
77) Bromobenzene	12.982	156	248005	54.743	ug/l		93
78) n-propylbenzene	13.041	91	1225582	57.836	ug/l		98
79) 2-Chlorotoluene	13.129	91	763590	56.407	ug/l		99
80) 1,3,5-Trimethylbenzene	13.176	105	843274	58.573	ug/l		100
81) trans-1,4-Dichloro-2-b...	12.741	75	113532	53.963	ug/l	#	86
82) 4-Chlorotoluene	13.223	91	770767	57.298	ug/l		98
83) tert-Butylbenzene	13.441	119	696142	59.072	ug/l		97
84) 1,2,4-Trimethylbenzene	13.488	105	857461	60.535	ug/l		99
85) sec-Butylbenzene	13.617	105	969354	59.370	ug/l		100
86) p-Isopropyltoluene	13.735	119	804085	60.405	ug/l		99
87) 1,3-Dichlorobenzene	13.735	146	453195	53.774	ug/l		99
88) 1,4-Dichlorobenzene	13.817	146	447008	52.255	ug/l		98
89) n-Butylbenzene	14.059	91	694122	60.541	ug/l		99
90) Hexachloroethane	14.335	117	147777	56.055	ug/l		87
91) 1,2-Dichlorobenzene	14.112	146	435861	54.785	ug/l		99
92) 1,2-Dibromo-3-Chloropr...	14.723	75	67515	60.682	ug/l		97
93) 1,2,4-Trichlorobenzene	15.394	180	215046	56.846	ug/l		98
94) Hexachlorobutadiene	15.506	225	94991	55.392	ug/l		99
95) Naphthalene	15.647	128	718902	56.038	ug/l		100
96) 1,2,3-Trichlorobenzene	15.847	180	206072	55.573	ug/l		97

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110723\
Data File : VN079959.D
Acq On : 07 Nov 2023 12:33
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Quant Time: Nov 08 04:25:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :

Manual Integrations
APPROVED

Reviewed By :John
Carlone
11/08/2023

Supervised By :Mahesh
Dadoda
11/08/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110723\
Data File : VN079959.D
Acq On : 07 Nov 2023 12:33
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :

Quant Time: Nov 08 04:25:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John


Carlone


11/08/2023

Supervised By :Mahesh

Dadoda


11/08/2023





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1067
Delta R.T. 0.001 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
Client Sample Id :






Tgt Ion: 168 Resp: 33104
Ion Ratio Lower Upper
168 100
99 61.8 53.0 79.6

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#2
Dichlorodifluoromethane
Concen: 48.420 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 85 Resp: 193075
Ion Ratio Lower Upper
85 100
87 32.0 15.8 47.4





#3
Chloromethane
Concen: 45.205 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 50 Resp: 225362
Ion Ratio Lower Upper
50 100
52 33.7 26.5 39.7

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#4
Vinyl Chloride
Concen: 54.003 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 62 Resp: 281870
Ion Ratio Lower Upper
62 100
64 32.9 26.2 39.2





#5
Bromomethane
Concen: 49.495 ug/l
RT: 2.936 min Scan# 10
Delta R.T. -0.018 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 94 Resp: 15696
Ion Ratio Lower Upper
94 100
96 94.9 73.1 109.7

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#6
Chloroethane
Concen: 42.009 ug/l
RT: 3.112 min Scan# 197
Delta R.T. -0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 64 Resp: 173256
Ion Ratio Lower Upper
64 100
66 31.9 25.4 38.2





#7
Trichlorofluoromethane
Concen: 52.184 ug/l
RT: 3.495 min Scan# 20
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:101 Resp: 338776
Ion Ratio Lower Upper
101 100
103 63.9 50.7 76.1

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#8
Diethyl Ether
Concen: 52.458 ug/l
RT: 3.959 min Scan# 341
Delta R.T. -0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 74 Resp: 129233
Ion Ratio Lower Upper
74 100
45 94.9 34.4 103.3





#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.909 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:101 Resp: 191590
Ion Ratio Lower Upper
101 100
85 45.0 36.6 55.0
151 72.3 57.6 86.4

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#10
Methyl Iodide
Concen: 52.872 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion:142 Resp: 227226
Ion Ratio Lower Upper
142 100
127 46.2 37.6 56.4
141 14.5 11.4 17.0





#11
Tert butyl alcohol
Concen: 279.111 ug/l
RT: 5.524 min Scan# 607
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion: 59 Resp: 179699
Ion Ratio Lower Upper
59 100
57 11.1 8.2 12.4

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#12
1,1-Dichloroethene
Concen: 50.208 ug/l
RT: 4.342 min Scan# 406
Delta R.T. -0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 96 Resp: 178907
Ion Ratio Lower Upper
96 100
61 172.8 112.7 169.1#
98 63.2 51.2 76.8





#13
Acrolein
Concen: 292.927 ug/l
RT: 4.183 min Scan# 31
Delta R.T. -0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 56 Resp: 142680
Ion Ratio Lower Upper
56 100
55 72.6 54.5 81.7

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#14
Allyl chloride
Concen: 54.275 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 41 Resp: 296211
Ion Ratio Lower Upper
41 100
39 79.7 58.8 88.2
76 39.6 39.1 58.7





#15
Acrylonitrile
Concen: 284.715 ug/l
RT: 5.718 min Scan# 641
Delta R.T. -0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 53 Resp: 571249
Ion Ratio Lower Upper
53 100
52 81.6 65.3 97.9
51 36.2 28.9 43.3

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#16
Acetone
Concen: 298.802 ug/l
RT: 4.424 min Scan# 420
Delta R.T. -0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 43 Resp: 529354
Ion Ratio Lower Upper
43 100
58 31.1 28.0 42.0





#17
Carbon Disulfide
Concen: 45.064 ug/l
RT: 4.706 min Scan# 469
Delta R.T. -0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 76 Resp: 46318
Ion Ratio Lower Upper
76 100
78 8.7 7.4 11.0

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#18
Methyl Acetate
Concen: 56.958 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 43 Resp: 482763
Ion Ratio Lower Upper
43 100
74 24.0 26.2 39.4#





#19
Methyl tert-butyl Ether
Concen: 56.416 ug/l
RT: 5.794 min Scan# 61
Delta R.T. -0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 73 Resp: 69211
Ion Ratio Lower Upper
73 100
57 21.8 15.6 23.4

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#20
Methylene Chloride
Concen: 49.113 ug/l
RT: 5.277 min Scan# 565
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 84 Resp: 219543
Ion Ratio Lower Upper
84 100
49 127.8 79.7 119.5#
51 39.3 25.2 37.8#
86 64.4 53.3 79.9





#21
trans-1,2-Dichloroethene
Concen: 50.325 ug/l
RT: 5.789 min Scan# 61
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 96 Resp: 19999
Ion Ratio Lower Upper
96 100
61 143.3 90.1 135.1
98 64.5 46.4 69.6

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#22
Diisopropyl ether
Concen: 57.759 ug/l
RT: 6.671 min Scan# 802
Delta R.T. -0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 45 Resp: 730519
Ion Ratio Lower Upper
45 100
43 53.3 45.5 68.3
87 27.3 31.8 47.8
59 12.4 11.6 17.4





#23
 Vinyl Acetate
 Concen: 294.961 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.000 min
 Lab File: VN079959.D
 Acq: 07 Nov 2023 12:33

Instrument :
 MSVOA_N
 ClientSampleId :






Tgt Ion: 43 Resp: 2057770
 Ion Ratio Lower Upper
 43 100
 86 10.8 13.0 19.6

Manual Integrations
 APPROVED




Reviewed By :John
 Carlone
 11/08/2023


Supervised By :Mahesh
 Dadoda
 11/08/2023



#24
 1,1-Dichloroethane
 Concen: 54.477 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. 0.000 min
 Lab File: VN079959.D
 Acq: 07 Nov 2023 12:33


Tgt Ion: 63 Resp: 418782
 Ion Ratio Lower Upper
 63 100
 98 5.7 3.9 11.7
 100 4.4 2.5 7.4





#25
2-Butanone
Concen: 305.465 ug/l
RT: 7.483 min Scan# 940
Delta R.T. -0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 43 Resp: 792449
Ion Ratio Lower Upper
43 100
72 25.9 30.2 45.2

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#26
2,2-Dichloropropane
Concen: 56.740 ug/l
RT: 7.494 min Scan# 942
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 77 Resp: 360730
Ion Ratio Lower Upper
77 100
97 21.9 11.2 33.6





#27
 cis-1,2-Dichloroethene
 Concen: 52.224 ug/l
 RT: 7.488 min Scan# 941
 Delta R.T. 0.000 min
 Lab File: VN079959.D
 Acq: 07 Nov 2023 12:33

Instrument :
 MSVOA_N
 ClientSampleId :






Tgt Ion: 96 Resp: 24826
 Ion Ratio Lower Upper
 96 100
 61 151.9 0.0 255.2
 98 64.5 0.0 128.8

Manual Integrations
 APPROVED




Reviewed By :John
 Carlone
 11/08/2023


Supervised By :Mahesh
 Dadoda
 11/08/2023



#28
 Bromochloromethane
 Concen: 48.592 ug/l
 RT: 7.818 min Scan# 997
 Delta R.T. 0.006 min
 Lab File: VN079959.D
 Acq: 07 Nov 2023 12:33

Tgt Ion: 49 Resp: 182041
 Ion Ratio Lower Upper
 49 100
 129 1.7 0.0 4.8
 130 70.8 75.3 112.9#





#29

Tetrahydrofuran

Concen: 300.137 ug/l

RT: 7.835 min Scan# 1000

Delta R.T. -0.006 min


Lab File: VN079959.D

Acq: 07 Nov 2023 12:33

Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion: 42 Resp: 50658

Ion Ratio Lower Upper

42 100

72 43.8 49.8 74.6

71 40.6 47.4 71.2

Manual Integrations

APPROVED

Reviewed By :John




Carlone

11/08/2023

Supervised By :Mahesh

Dadoda

11/08/2023



#30

Chloroform

Concen: 51.787 ug/l

RT: 7.965 min Scan# 1022

Delta R.T. -0.006 min

Lab File: VN079959.D




Acq: 07 Nov 2023 12:33


Tgt Ion: 83 Resp: 432428

Ion Ratio Lower Upper

83 100


85 61.5 55.4 83.0





#31
Cyclohexane
Concen: 49.686 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 56 Resp: 331135
Ion Ratio Lower Upper
56 100
69 31.6 26.8 40.2
84 88.7 81.8 122.6

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#32
1,1,1-Trichloroethane
Concen: 54.979 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 97 Resp: 374066
Ion Ratio Lower Upper
97 100
99 64.7 50.3 75.5
61 49.4 34.7 52.1





#33
1,2-Dichloroethane-d4
Concen: 56.788 ug/l
RT: 8.583 min Scan# 1127
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 65 Resp: 28641
Ion Ratio Lower Upper
65 100
67 51.9 0.0 106.6

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.106 min Scan# 1216
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion:114 Resp: 567705
Ion Ratio Lower Upper
114 100
63 22.3 0.0 36.4
88 16.3 0.0 30.6





#35
Dibromofluoromethane
Concen: 54.879 ug/l
RT: 8.171 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 113 Resp: 22050
Ion Ratio Lower Upper
113 100
111 103.5 82.6 123.8
192 18.0 12.7 19.1

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#36
1,1-Dichloropropene
Concen: 54.866 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. -0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 75 Resp: 314768
Ion Ratio Lower Upper
75 100
110 32.9 19.1 57.3
77 30.8 25.0 37.4





#37
Ethyl Acetate
Concen: 60.108 ug/l
RT: 7.565 min Scan# 91
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 43 Resp: 316262
Ion Ratio Lower Upper
43 100
61 13.3 13.2 19.8
70 10.4 11.3 16.9

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#38
Carbon Tetrachloride
Concen: 55.215 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion:117 Resp: 319248
Ion Ratio Lower Upper
117 100
119 92.8 77.3 115.9
121 30.0 23.3 34.9





#39
Methylcyclohexane
Concen: 55.795 ug/l
RT: 9.606 min Scan# 1130
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 83 Resp: 301238
Ion Ratio Lower Upper
83 100
55 78.2 49.1 73.7
98 49.2 40.6 61.0

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#40
Benzene
Concen: 54.104 ug/l
RT: 8.612 min Scan# 1132
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 78 Resp: 929678
Ion Ratio Lower Upper
78 100
77 23.5 19.0 28.6





#41
Methacrylonitrile
Concen: 62.046 ug/l
RT: 7.783 min Scan# 991
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 41 Resp: 173350
Ion Ratio Lower Upper
41 100
39 58.3 42.6 64.0
67 73.5 77.5 116.3
52 31.0 28.9 43.3

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#42
1,2-Dichloroethane
Concen: 57.910 ug/l
RT: 8.671 min Scan# 1142
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 62 Resp: 347589
Ion Ratio Lower Upper
62 100
98 8.8 0.0 19.8





#43
Isopropyl Acetate
Concen: 59.454 ug/l
RT: 8.694 min Scan# 1146
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 43 Resp: 548580
Ion Ratio Lower Upper
43 100
61 25.0 24.7 37.1
87 12.6 14.6 21.8

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#44
Trichloroethene
Concen: 51.459 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Tgt Ion:130 Resp: 219415
Ion Ratio Lower Upper
130 100
95 107.2 0.0 204.6





#45

1,2-Dichloropropane

Concen: 57.830 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.000 min


Lab File: VN079959.D

Acq: 07 Nov 2023 12:33

Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion: 63 Resp: 253949

Ion Ratio Lower Upper

63 100

65 31.5 23.9 35.9

Manual Integrations

APPROVED

Reviewed By :John




Carlone

11/08/2023

Supervised By :Mahesh

Dadoda

11/08/2023



#46

Dibromomethane

Concen: 54.780 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.000 min

Lab File: VN079959.D

Acq: 07 Nov 2023 12:33




Tgt Ion: 93 Resp: 164253


Ion Ratio Lower Upper

93 100

95 83.3 63.8 95.8


174 88.2 65.6 98.4





#47
Bromodichloromethane
Concen: 54.052 ug/l
RT: 9.894 min Scan# 1350
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 83 Resp: 34564
Ion Ratio Lower Upper
83 100
85 62.3 49.6 74.4
127 7.9 7.0 10.6

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#48
Methyl methacrylate
Concen: 63.939 ug/l
RT: 9.682 min Scan# 1314
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Tgt Ion: 41 Resp: 258415
Ion Ratio Lower Upper
41 100
69 83.7 90.4 135.6#
39 59.4 54.6 81.8





#49

1,4-Dioxane

Concen: 1244.214 ug/l

RT: 9.700 min Scan# 11

Delta R.T. 0.006 min


Lab File: VN079959.D

Acq: 07 Nov 2023 12:33

Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion:	88	Resp:	8885
Ion	Ratio	Lower	Upper
88	100		
43	31.1	17.0	25.6
58	71.3	47.0	70.4

Manual Integrations
APPROVED

Reviewed By :John




Carlone

11/08/2023

Supervised By :Mahesh

Dadoda

11/08/2023



#50

Toluene-d8

Concen: 55.523 ug/l




RT: 10.571 min Scan# 1465


Delta R.T. 0.000 min

Lab File: VN079959.D

Acq: 07 Nov 2023 12:33


Tgt Ion:	98	Resp:	821272
Ion	Ratio	Lower	Upper
98	100		
100	65.3	52.0	78.0





#51
4-Methyl-2-Pentanone
Concen: 318.345 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 43 Resp: 165373
Ion Ratio Lower Upper
43 100
58 39.5 36.6 54.8

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#52
Toluene
Concen: 56.468 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 92 Resp: 575230
Ion Ratio Lower Upper
92 100
91 173.4 137.4 206.0





#53
t-1,3-Dichloropropene
Concen: 60.670 ug/l
RT: 10.841 min Scan# 1511
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 75 Resp: 38033
Ion Ratio Lower Upper
75 100
77 30.1 25.4 38.0

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#54
cis-1,3-Dichloropropene
Concen: 58.678 ug/l
RT: 10.318 min Scan# 1422
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 75 Resp: 400939
Ion Ratio Lower Upper
75 100
77 31.1 26.1 39.1
39 49.8 31.6 47.4#





#55
1,1,2-Trichloroethane
Concen: 56.605 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion: 97 Resp: 23193

Ion	Ratio	Lower	Upper
97	100		
83	87.8	65.1	97.7
85	55.0	43.4	65.2
99	62.2	49.2	73.8

Manual Integrations
APPROVED

Reviewed By :John
Carlone
11/08/2023




Supervised By :Mahesh
Dadoda
11/08/2023




#56
Ethyl methacrylate
Concen: 63.955 ug/l
RT: 10.876 min Scan# 1517
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Tgt Ion: 69 Resp: 373782

Ion	Ratio	Lower	Upper
69	100		
41	65.9	39.6	59.4
39	36.7	28.6	43.0





#57

1,3-Dichloropropane

Concen: 57.553 ug/l

RT: 11.165 min Scan# 1566

Delta R.T. 0.000 min


Lab File: VN079959.D

Acq: 07 Nov 2023 12:33

Instrument :

MSVOA_N

ClientSampleId :



Tgt Ion: 76 Resp: 415660

Ion Ratio Lower Upper

76 100

78 31.9 26.1 39.1

Manual Integrations

APPROVED

Reviewed By :John




Carlone

11/08/2023

Supervised By :Mahesh

Dadoda

11/08/2023



#58

2-Chloroethyl Vinyl ether

Concen: 290.350 ug/l

RT: 10.165 min Scan# 1396

Delta R.T. 0.000 min

Lab File: VN079959.D




Acq: 07 Nov 2023 12:33


Tgt Ion: 63 Resp: 860068

Ion Ratio Lower Upper

63 100


106 27.1 26.2 39.4





#59
2-Hexanone
Concen: 306.860 ug/l
RT: 11.200 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 43 Resp: 121946
Ion Ratio Lower Upper
43 100
58 55.2 31.7 95.1

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#60
Dibromochloromethane
Concen: 56.586 ug/l
RT: 11.365 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion:129 Resp: 248177
Ion Ratio Lower Upper
129 100
127 77.1 38.6 116.0





#61
1,2-Dibromoethane
Concen: 56.177 ug/l
RT: 11.476 min Scan# 1618
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:107 Resp: 231024
Ion Ratio Lower Upper
107 100
109 93.5 76.6 114.8

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#62
4-Bromofluorobenzene
Concen: 61.602 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 95 Resp: 302200
Ion Ratio Lower Upper
95 100
174 70.6 0.0 143.8
176 69.3 0.0 135.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:117 Resp: 515160
Ion Ratio Lower Upper
117 100
82 58.1 44.0 66.0
119 32.0 25.6 38.4

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#64
Tetrachloroethene
Concen: 51.613 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion:164 Resp: 172207
Ion Ratio Lower Upper
164 100
166 128.7 101.4 152.0
129 98.7 82.4 123.6
131 94.2 78.2 117.4





#65
Chlorobenzene
Concen: 54.490 ug/l
RT: 11.894 min Scan# 1690
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :





Tgt Ion:112 Resp: 608925
Ion Ratio Lower Upper
112 100
114 31.9 25.0 37.4

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#66
1,1,1,2-Tetrachloroethane
Concen: 57.385 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion:131 Resp: 231045
Ion Ratio Lower Upper
131 100
133 96.0 46.4 139.1
119 65.1 34.1 102.2





#67
Ethyl Benzene
Concen: 57.729 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 91 Resp: 112396
Ion Ratio Lower Upper
91 100
106 29.4 25.6 38.4

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#68
m/p-Xylenes
Concen: 115.516 ug/l
RT: 12.076 min Scan# 1721
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion:106 Resp: 837812
Ion Ratio Lower Upper
106 100
91 211.3 159.6 239.4





#69
o-Xylene
Concen: 58.357 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:106 Resp: 40756
Ion Ratio Lower Upper
106 100
91 223.9 105.2 315.5

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#70
Styrene
Concen: 61.297 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion:104 Resp: 687988
Ion Ratio Lower Upper
104 100
78 52.4 39.5 59.3
103 55.1 43.2 64.8





#71
Bromoform
Concen: 60.365 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:173 Resp: 17258
Ion Ratio Lower Upper
173 100
175 48.0 24.3 73.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion:152 Resp: 243820
Ion Ratio Lower Upper
152 100
115 63.6 32.3 96.8
150 174.6 0.0 355.8





#73
Isopropylbenzene
Concen: 56.203 ug/l
RT: 12.700 min Scan# 1827
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:105 Resp: 104000
Ion Ratio Lower Upper
105 100
120 26.0 13.1 39.1

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#74
N-ethyl acetate
Concen: 57.121 ug/l
RT: 12.500 min Scan# 1793
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 43 Resp: 461363
Ion Ratio Lower Upper
43 100
70 43.7 43.4 65.2
55 25.2 21.4 32.0
61 24.9 22.8 34.2





#75
1,1,2,2-Tetrachloroethane
Concen: 53.703 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 83 Resp: 352338
Ion Ratio Lower Upper
83 100
131 9.8 5.0 15.0
85 64.7 31.9 95.7

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#76
1,2,3-Trichloropropane
Concen: 49.810 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 75 Resp: 285121
Ion Ratio Lower Upper
75 100
77 200.3 84.5 253.5





#77
Bromobenzene
Concen: 54.743 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:156 Resp: 248000
Ion Ratio Lower Upper
156 100
77 230.3 107.7 323.3
158 95.6 48.4 145.3

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#78
n-propylbenzene
Concen: 57.836 ug/l
RT: 13.041 min Scan# 1885
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Tgt Ion: 91 Resp: 1225582
Ion Ratio Lower Upper
91 100
120 21.9 11.5 34.4





#79

2-Chlorotoluene

Concen: 56.407 ug/l

RT: 13.129 min Scan# 1900

Delta R.T. 0.000 min

Lab File: VN079959.D

Acq: 07 Nov 2023 12:33

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion: 91 Resp: 763590

Ion Ratio Lower Upper

91 100

126 31.9 16.3 48.8

Manual Integrations

APPROVED

Reviewed By :John





Carlone

11/08/2023

Supervised By :Mahesh

Dadoda

11/08/2023



#80

1,3,5-Trimethylbenzene

Concen: 58.573 ug/l

RT: 13.176 min Scan# 1908

Delta R.T. 0.000 min

Lab File: VN079959.D




Acq: 07 Nov 2023 12:33


Tgt Ion:105 Resp: 843274

Ion Ratio Lower Upper

105 100


120 48.4 24.1 72.3





#81
trans-1,4-Dichloro-2-butene
Concen: 53.963 ug/l
RT: 12.741 min Scan# 11353
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion: 75 Resp: 11353
Ion Ratio Lower Upper
75 100
53 103.9 67.4 101.2
89 47.2 38.8 58.2

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#82
4-Chlorotoluene
Concen: 57.298 ug/l
RT: 13.223 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 91 Resp: 770767
Ion Ratio Lower Upper
91 100
126 31.9 16.7 50.0





#83
tert-Butylbenzene
Concen: 59.072 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:119 Resp: 696142
Ion Ratio Lower Upper
119 100
91 70.4 33.9 101.7
134 27.7 13.2 39.6

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#84
1,2,4-Trimethylbenzene
Concen: 60.535 ug/l
RT: 13.488 min Scan# 1961
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion:105 Resp: 857461
Ion Ratio Lower Upper
105 100
120 43.5 22.0 66.0





#85
 sec-Butylbenzene
 Concen: 59.370 ug/l
 RT: 13.617 min Scan# 1983
 Delta R.T. 0.000 min
 Lab File: VN079959.D
 Acq: 07 Nov 2023 12:33

Instrument :
 MSVOA_N
 ClientSampleId :






Tgt Ion:105 Resp: 969354
 Ion Ratio Lower Upper
 105 100
 134 19.6 9.8 29.3

Manual Integrations
 APPROVED




Reviewed By :John
 Carlone
 11/08/2023


Supervised By :Mahesh
 Dadoda
 11/08/2023



#86
 p-Isopropyltoluene
 Concen: 60.405 ug/l
 RT: 13.735 min Scan# 2003
 Delta R.T. 0.006 min
 Lab File: VN079959.D
 Acq: 07 Nov 2023 12:33


Tgt Ion:119 Resp: 804085
 Ion Ratio Lower Upper
 119 100
 134 25.7 12.6 37.8
 91 24.8 12.4 37.2





#87
1,3-Dichlorobenzene
Concen: 53.774 ug/l
RT: 13.735 min Scan# 2003
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:146 Resp: 45319
Ion Ratio Lower Upper
146 100
111 43.1 22.1 66.3
148 63.3 32.0 96.0

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#88
1,4-Dichlorobenzene
Concen: 52.255 ug/l
RT: 13.817 min Scan# 2017
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Tgt Ion:146 Resp: 447008
Ion Ratio Lower Upper
146 100
111 42.4 20.6 61.8
148 64.4 31.6 94.8





#89

n-Butylbenzene

Concen: 60.541 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. 0.000 min

Lab File: VN079959.D

Acq: 07 Nov 2023 12:33

Instrument :

MSVOA_N

ClientSampleId :

Tgt Ion:	91	Resp:	69412
Ion Ratio	Lower	Upper	
91	100		
92	52.2	26.4	79.0
134	23.9	12.6	37.6

Manual Integrations

APPROVED

Reviewed By :John





Carlone

11/08/2023

Supervised By :Mahesh

Dadoda

11/08/2023



#90

Hexachloroethane

Concen: 56.055 ug/l




RT: 14.335 min Scan# 2105


Delta R.T. 0.000 min

Lab File: VN079959.D

Acq: 07 Nov 2023 12:33


Tgt Ion:	117	Resp:	147777
Ion Ratio	Lower	Upper	
117	100		
201	68.1	29.3	88.0





#91
1,2-Dichlorobenzene
Concen: 54.785 ug/l
RT: 14.112 min Scan# 2066
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:146 Resp: 43586
Ion Ratio Lower Upper
146 100
111 45.0 22.0 66.0
148 63.7 31.8 95.3

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#92
1,2-Dibromo-3-Chloropropane
Concen: 60.682 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion: 75 Resp: 67515
Ion Ratio Lower Upper
75 100
155 75.3 39.1 117.2
157 97.4 50.3 150.9





#93
1,2,4-Trichlorobenzene
Concen: 56.846 ug/l
RT: 15.394 min Scan# 2285
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:180 Resp: 215040
Ion Ratio Lower Upper
180 100
182 94.5 46.1 138.2
145 32.1 16.6 49.8

Manual Integrations
APPROVED




Reviewed By :John
Carlone
11/08/2023


Supervised By :Mahesh
Dadoda
11/08/2023



#94
Hexachlorobutadiene
Concen: 55.392 ug/l
RT: 15.506 min Scan# 2304
Delta R.T. 0.000 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33


Tgt Ion:225 Resp: 94991
Ion Ratio Lower Upper
225 100
223 62.8 31.3 93.8
227 65.2 31.9 95.9





#95
Naphthalene
Concen: 56.038 ug/l
RT: 15.647 min Scan# 2327
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33

Instrument :
MSVOA_N
ClientSampleId :






Tgt Ion:128 Resp: 718902
Ion Ratio Lower Upper
128 100
127 12.8 10.0 15.0
129 11.0 8.7 13.1


Manual Integrations
APPROVED

Reviewed By :John
Carlone
11/08/2023

Supervised By :Mahesh
Dadoda
11/08/2023



#96
1,2,3-Trichlorobenzene
Concen: 55.573 ug/l
RT: 15.847 min Scan# 2362
Delta R.T. 0.006 min
Lab File: VN079959.D
Acq: 07 Nov 2023 12:33



Tgt Ion:180 Resp: 206072
Ion Ratio Lower Upper
180 100
182 96.0 46.2 138.6
145 33.9 17.5 52.6

