

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110818\
 Data File : VN052221.D
 Acq On : 8 Nov 2018 15:33
 Operator : MD\SY
 Sample : J5878-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SPDES-1

Quant Time: Nov 08 15:54:49 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA N\METHODS\624N110118W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Fri Nov 02 05:39:38 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	7.19	128	59719	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.58	114	347893	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.41	117	293639	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.02	65	131253	29.61	ug/l	0.00
Spiked Amount	30.000	Range	50 - 169	Recovery	=	98.70%
60) 4-Bromofluorobenzene	12.40	95	124095	26.42	ug/l	0.00
Spiked Amount	30.000	Range	56 - 143	Recovery	=	88.07%
63) Toluene-d8	10.09	98	417580	30.24	ug/l	0.00
Spiked Amount	30.000	Range	66 - 137	Recovery	=	100.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110818\
Data File : VN052221.D
Acq On : 8 Nov 2018 15:33
Operator : MD\SY
Sample : J5878-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SPDES-1

Quant Time: Nov 08 15:54:49 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\624N110118W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Fri Nov 02 05:39:38 2018
Response via : Initial Calibration

