

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110822\
 Data File : VN075202.D
 Acq On : 08 Nov 2022 11:19
 Operator : JC\MD
 Sample : VN1108MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1108MBL01

Quant Time: Nov 09 01:29:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	96452	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	179637	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	151543	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	61154	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	70257	54.061	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.120%	
35) Dibromofluoromethane	8.171	113	58018	55.016	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.040%	
50) Toluene-d8	10.565	98	201462	49.240	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 98.480%	
62) 4-Bromofluorobenzene	12.853	95	64616	44.959	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.920%	

Target Compounds Qvalue

 (#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110822\
Data File : VN075202.D
Acq On : 08 Nov 2022 11:19
Operator : JC\MD
Sample : VN1108MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1108MBL01

Quant Time: Nov 09 01:29:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

