

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110822\
 Data File : VN075201.D
 Acq On : 08 Nov 2022 10:07
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 09 01:24:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 T	Dichlorodifluoromethane	0.650	0.530	18.5	83	0.00
3 P	Chloromethane	0.691	0.527	23.7	79	0.00
4 C	Vinyl Chloride	0.986	0.801	18.8#	81	0.00
5 T	Bromomethane	0.808	0.749	7.3	95	0.00
6 T	Chloroethane	0.707	0.587	17.0	86	0.00
7 T	Trichlorofluoromethane	1.007	0.975	3.2	98	0.00
8 T	Diethyl Ether	0.367	0.348	5.2	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.556	0.572	-2.9	105	0.00
10 T	Methyl Iodide	0.828	0.758	8.5	96	0.00
11 T	Tert butyl alcohol	0.157	0.131	16.6	90	0.00
12 CM	1,1-Dichloroethene	0.530	0.507	4.3#	97	0.00
13 T	Acrolein	0.129	0.114	11.6	92	0.00
14 T	Allyl chloride	0.774	0.729	5.8	92	0.00
15 T	Acrylonitrile	0.332	0.313	5.7	96	0.00
16 T	Acetone	0.285	0.313	-9.8	117	0.00
17 T	Carbon Disulfide	1.333	1.124	15.7	84	0.00
18 T	Methyl Acetate	0.932	0.924	0.9	102	0.00
19 T	Methyl tert-butyl Ether	1.970	1.934	1.8	98	0.00
20 T	Methylene Chloride	0.636	0.607	4.6	99	0.00
21 T	trans-1,2-Dichloroethene	0.580	0.554	4.5	98	0.00
22 T	Diisopropyl ether	1.708	1.692	0.9	99	0.00
23 T	Vinyl Acetate	1.427	1.346	5.7	93	0.00
24 P	1,1-Dichloroethane	1.080	1.050	2.8	99	0.00
25 T	2-Butanone	0.449	0.440	2.0	101	0.00
26 T	2,2-Dichloropropane	0.994	0.992	0.2	101	0.00
27 T	cis-1,2-Dichloroethene	0.693	0.693	0.0	101	0.00
28 T	Bromochloromethane	0.455	0.432	5.1	94	0.00
29 T	Tetrahydrofuran	0.290	0.272	6.2	94	0.00
30 C	Chloroform	1.173	1.214	-3.5#	104	0.00
31 T	Cyclohexane	1.002	0.896	10.6	92	0.00
32 T	1,1,1-Trichloroethane	1.065	1.093	-2.6	101	0.00
33 S	1,2-Dichloroethane-d4	0.674	0.671	0.4	103	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	95	0.00
35 S	Dibromofluoromethane	0.294	0.319	-8.5	107	0.00
36 T	1,1-Dichloropropene	0.454	0.470	-3.5	99	0.00
37 T	Ethyl Acetate	0.498	0.457	8.2	94	0.00
38 T	Carbon Tetrachloride	0.500	0.536	-7.2	101	0.00
39 T	Methylcyclohexane	0.600	0.611	-1.8	97	0.00
40 TM	Benzene	1.391	1.440	-3.5	100	0.00
41 T	Methacrylonitrile	0.239	0.240	-0.4	97	0.00
42 TM	1,2-Dichloroethane	0.506	0.507	-0.2	97	0.00
43 T	Isopropyl Acetate	0.773	0.735	4.9	91	0.00
44 TM	Trichloroethene	0.349	0.367	-5.2	104	0.00
45 C	1,2-Dichloropropane	0.331	0.355	-7.3#	101	0.00
46 T	Dibromomethane	0.258	0.266	-3.1	101	0.00
47 T	Bromodichloromethane	0.500	0.525	-5.0	102	0.00
48 T	Methyl methacrylate	0.355	0.350	1.4	98	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110822\
 Data File : VN075201.D
 Acq On : 08 Nov 2022 10:07
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 09 01:24:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.010	-11.1	107	0.00
50 S	Toluene-d8	1.139	1.200	-5.4	102	0.00
51 T	4-Methyl-2-Pentanone	0.499	0.487	2.4	95	0.00
52 CM	Toluene	0.930	0.989	-6.3#	98	0.00
53 T	t-1,3-Dichloropropene	0.527	0.553	-4.9	96	0.00
54 T	cis-1,3-Dichloropropene	0.559	0.587	-5.0	98	0.00
55 T	1,1,2-Trichloroethane	0.369	0.386	-4.6	99	0.00
56 T	Ethyl methacrylate	0.563	0.573	-1.8	94	0.00
57 T	1,3-Dichloropropane	0.614	0.654	-6.5	102	0.00
58 T	2-Chloroethyl Vinyl ether	0.217	0.207	4.6	86	0.00
59 T	2-Hexanone	0.379	0.367	3.2	94	0.00
60 T	Dibromochloromethane	0.389	0.424	-9.0	100	0.00
61 T	1,2-Dibromoethane	0.378	0.404	-6.9	99	0.00
62 S	4-Bromofluorobenzene	0.400	0.439	-9.7	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	93	0.00
64 T	Tetrachloroethene	0.307	0.346	-12.7	110	0.00
65 PM	Chlorobenzene	1.067	1.101	-3.2	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.396	0.432	-9.1	102	0.00
67 C	Ethyl Benzene	1.920	2.054	-7.0#	98	0.00
68 T	m/p-Xylenes	0.770	0.830	-7.8	96	0.00
69 T	o-Xylene	0.762	0.807	-5.9	97	0.00
70 T	Styrene	1.215	1.333	-9.7	98	0.00
71 P	Bromoform	0.296	0.324	-9.5	100	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	92	0.00
73 T	Isopropylbenzene	4.014	4.025	-0.3	96	0.00
74 T	N-amyl acetate	1.485	1.365	8.1	90	0.00
75 P	1,1,2,2-Tetrachloroethane	1.309	1.258	3.9	96	0.00
76 T	1,2,3-Trichloropropane	1.062	1.081	-1.8	93	0.00
77 T	Bromobenzene	0.884	0.915	-3.5	101	0.00
78 T	n-propylbenzene	4.507	4.697	-4.2	98	0.00
79 T	2-Chlorotoluene	2.754	2.749	0.2	95	0.00
80 T	1,3,5-Trimethylbenzene	3.371	3.484	-3.4	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.414	0.390	5.8	89	0.00
82 T	4-Chlorotoluene	2.754	2.749	0.2	95	0.00
83 T	tert-Butylbenzene	2.996	3.073	-2.6	97	0.00
84 T	1,2,4-Trimethylbenzene	3.411	3.588	-5.2	98	0.00
85 T	sec-Butylbenzene	4.191	4.467	-6.6	99	0.00
86 T	p-Isopropyltoluene	3.461	3.798	-9.7	99	0.00
87 T	1,3-Dichlorobenzene	1.731	1.892	-9.3	103	0.00
88 T	1,4-Dichlorobenzene	1.713	1.861	-8.6	103	0.00
89 T	n-Butylbenzene	2.787	3.138	-12.6	100	0.00
90 T	Hexachloroethane	0.635	0.660	-3.9	97	0.00
91 T	1,2-Dichlorobenzene	1.720	1.827	-6.2	100	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.271	0.254	6.3	89	0.00
93 T	1,2,4-Trichlorobenzene	0.740	0.864	-16.8	105	0.00
94 T	Hexachlorobutadiene	0.361	0.405	-12.2	110	0.00
95 T	Naphthalene	2.847	2.996	-5.2	94	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110822\
Data File : VN075201.D
Acq On : 08 Nov 2022 10:07
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 09 01:24:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.736	0.843	-14.5	105	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6