

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
 Data File : VN084746.D
 Acq On : 08 Nov 2024 13:44
 Operator : JC\MD
 Sample : P4774-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-190-TB-20241105

Quant Time: Nov 08 22:30:13 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	164125	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	293667	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	256546	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	111909	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	117110	49.403	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	98.800%		
35) Dibromofluoromethane	8.165	113	96422	48.508	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.020%		
50) Toluene-d8	10.565	98	332676	45.433	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.860%		
62) 4-Bromofluorobenzene	12.847	95	125399	45.823	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	91.640%		
Target Compounds						
18) Methyl Acetate	5.024	43	16971	3.035	ug/l	Qvalue 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
Data File : VN084746.D
Acq On : 08 Nov 2024 13:44
Operator : JC\MD
Sample : P4774-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-190-TB-20241105

Quant Time: Nov 08 22:30:13 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

