

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
 Data File : VN084748.D
 Acq On : 08 Nov 2024 14:32
 Operator : JC\MD
 Sample : P4774-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-190-EB-20241106

Quant Time: Nov 08 22:31:00 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	172082	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	302375	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	260963	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	112988	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	121482	48.877	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.760%		
35) Dibromofluoromethane	8.159	113	100718	49.210	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.420%		
50) Toluene-d8	10.565	98	344401	45.680	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	91.360%		
62) 4-Bromofluorobenzene	12.847	95	127539	45.263	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	90.520%		
Target Compounds						
29) Tetrahydrofuran	7.835	42	4345	5.648	ug/l	Qvalue 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
Data File : VN084748.D
Acq On : 08 Nov 2024 14:32
Operator : JC\MD
Sample : P4774-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-190-EB-20241106

Quant Time: Nov 08 22:31:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

