

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
 Data File : VN084749.D
 Acq On : 08 Nov 2024 14:56
 Operator : JC\MD
 Sample : P4774-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-190-DUP-20241105

Quant Time: Nov 08 22:31:29 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	176583	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	310479	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	273960	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	123463	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	126348	49.539	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.080%		
35) Dibromofluoromethane	8.165	113	102080	48.573	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.140%		
50) Toluene-d8	10.565	98	358668	46.331	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	92.660%		
62) 4-Bromofluorobenzene	12.847	95	135882	46.965	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	93.940%		
Target Compounds						
44) Trichloroethene	9.353	130	842	0.390	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
Data File : VN084749.D
Acq On : 08 Nov 2024 14:56
Operator : JC\MD
Sample : P4774-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-190-DUP-20241105

Quant Time: Nov 08 22:31:29 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

