

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
 Data File : VN084753.D
 Acq On : 08 Nov 2024 16:32
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Nov 08 22:33:22 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.218	168	172338	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	306054	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	265177	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	123726	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.571	65	123505	49.617	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.240%		
35) Dibromofluoromethane	8.159	113	100868	48.690	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.380%		
50) Toluene-d8	10.565	98	347789	45.575	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	91.140%		
62) 4-Bromofluorobenzene	12.847	95	132599	46.493	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	92.980%		
Target Compounds						
44) Trichloroethene	9.353	130	8051	3.784	ug/l	Qvalue 93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
Data File : VN084753.D
Acq On : 08 Nov 2024 16:32
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Nov 08 22:33:22 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

