

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
 Data File : VN084754.D
 Acq On : 08 Nov 2024 16:56
 Operator : JC\MD
 Sample : P4722-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1(3.5)

Quant Time: Nov 08 22:33:47 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	173136	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	304734	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	285280	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	132176	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	124399	49.746	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	99.500%		
35) Dibromofluoromethane	8.159	113	101220	49.072	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.140%		
50) Toluene-d8	10.565	98	368310	48.473	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	96.940%		
62) 4-Bromofluorobenzene	12.847	95	142767	50.275	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	100.560%		
Target Compounds						
16) Acetone	4.436	43	14853	17.914	ug/l	96
20) Methylene Chloride	5.265	84	3582	1.629	ug/l	91
25) 2-Butanone	7.488	43	8576	7.351	ug/l	96
43) Isopropyl Acetate	8.688	43	94549	19.122	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
Data File : VN084754.D
Acq On : 08 Nov 2024 16:56
Operator : JC\MD
Sample : P4722-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1(3.5)

Quant Time: Nov 08 22:33:47 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

