

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
 Data File : VN084756.D
 Acq On : 08 Nov 2024 17:44
 Operator : JC\MD
 Sample : P4739-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-14

Quant Time: Nov 08 22:34:55 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	177154	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	306226	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	280854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	129905	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.576	65	122911	48.036	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	96.080%		
35) Dibromofluoromethane	8.165	113	100390	48.433	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	96.860%		
50) Toluene-d8	10.565	98	370427	48.514	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	97.020%		
62) 4-Bromofluorobenzene	12.847	95	138147	48.411	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	96.820%		
Target Compounds						
16) Acetone	4.430	43	19921	24.378	ug/l	97
20) Methylene Chloride	5.265	84	3790	1.684	ug/l #	85
25) 2-Butanone	7.482	43	10492	8.789	ug/l	96
43) Isopropyl Acetate	8.688	43	121752	24.781	ug/l #	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
Data File : VN084756.D
Acq On : 08 Nov 2024 17:44
Operator : JC\MD
Sample : P4739-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-14

Quant Time: Nov 08 22:34:55 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

