

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
 Data File : VN084760.D
 Acq On : 08 Nov 2024 19:20
 Operator : JC\MD
 Sample : VIBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VIBLK

Quant Time: Nov 08 22:37:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 31 18:45:38 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	165582	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	289279	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	249939	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	113435	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	116870	48.868	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	97.740%		
35) Dibromofluoromethane	8.165	113	95416	48.730	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.460%		
50) Toluene-d8	10.565	98	337240	46.755	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	93.520%		
62) 4-Bromofluorobenzene	12.847	95	125273	46.471	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery =	92.940%		
Target Compounds						
67) Ethyl Benzene	11.965	91	11366	1.218	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110824\
Data File : VN084760.D
Acq On : 08 Nov 2024 19:20
Operator : JC\MD
Sample : VIBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VIBLK

Quant Time: Nov 08 22:37:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N103024W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 31 18:45:38 2024
Response via : Initial Calibration

