

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110918\
 Data File : VN052244.D
 Acq On : 9 Nov 2018 22:56
 Operator : MD\SY
 Sample : J5879-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 22 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 4626-22221

Quant Time: Nov 12 04:10:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 00:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	326076	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	551925	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	458361	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	164523	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	213262	47.82	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.64%	
35) Dibromofluoromethane	7.59	113	174412	48.79	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.58%	
50) Toluene-d8	10.09	98	677060	48.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.78%	
62) 4-Bromofluorobenzene	12.40	95	198327	43.83	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.66%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110918\
Data File : VN052244.D
Acq On : 9 Nov 2018 22:56
Operator : MD\SY
Sample : J5879-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 22 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
4626-22221

Quant Time: Nov 12 04:10:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 00:34:27 2018
Response via : Initial Calibration

