

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110918\
 Data File : VN052245.D
 Acq On : 9 Nov 2018 23:23
 Operator : MD\SY
 Sample : J5269-01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-12-GW

Quant Time: Nov 12 04:11:41 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 00:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	338396	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	560817	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	467341	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	160967	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	218425	47.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.38%	
35) Dibromofluoromethane	7.59	113	176424	48.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.14%	
50) Toluene-d8	10.09	98	687214	48.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.66%	
62) 4-Bromofluorobenzene	12.40	95	199882	43.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.94%	
Target Compounds						
52) Toluene	10.16	92	21201	2.169	ug/l	95
64) Tetrachloroethene	10.63	164	2866	0.876	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110918\
Data File : VN052245.D
Acq On : 9 Nov 2018 23:23
Operator : MD\SY
Sample : J5269-01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
SB-12-GW

Quant Time: Nov 12 04:11:41 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 00:34:27 2018
Response via : Initial Calibration

