

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN110918\
 Data File : VN052248.D
 Acq On : 10 Nov 2018 00:43
 Operator : MD\SY
 Sample : J5903-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 28

Instrument :
 MSVOA_N
 ClientSampleId :
 RAV-006

Quant Time: Nov 12 04:16:23 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
 Quant Title : SW846 8260
 QLast Update : Mon Nov 12 00:34:27 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	316262	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	529186	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	440443	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	142658	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	210098	48.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.14%	
35) Dibromofluoromethane	7.59	113	170389	49.71	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.42%	
50) Toluene-d8	10.09	98	653660	48.72	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.44%	
62) 4-Bromofluorobenzene	12.40	95	183549	42.31	ug/l	0.00
Spiked Amount	50.000		Recovery	=	84.62%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN110918\
Data File : VN052248.D
Acq On : 10 Nov 2018 00:43
Operator : MD\SY
Sample : J5903-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 28

Instrument :
MSVOA_N
ClientSampleId :
RAV-006

Quant Time: Nov 12 04:16:23 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110918W.M
Quant Title : SW846 8260
QLast Update : Mon Nov 12 00:34:27 2018
Response via : Initial Calibration

