

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110921\
 Data File : VN069423.D
 Acq On : 9 Nov 2021 17:37
 Operator : JC/MD
 Sample : VN1109WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1109WBL01

Quant Time: Nov 10 16:37:30 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	1246509	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.971	114	2218949	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	2176267	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	807678	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	941502	52.092	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	104.180%
35) Dibromofluoromethane	8.024	113	655208	49.233	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.460%
50) Toluene-d8	10.443	98	2846880	54.391	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	108.780%
62) 4-Bromofluorobenzene	12.734	95	972338	50.568	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110921\
Data File : VN069423.D
Acq On : 9 Nov 2021 17:37
Operator : JC/MD
Sample : VN1109WBL01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBL01

Quant Time: Nov 10 16:37:30 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

