

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110921\
 Data File : VN069430.D
 Acq On : 9 Nov 2021 20:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 10 16:20:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	95	0.00
2 T	Dichlorodifluoromethane	0.534	0.538	-0.7	91	0.00
3 P	Chloromethane	0.619	0.591	4.5	94	0.00
4 C	Vinyl Chloride	0.620	0.630	-1.6#	95	0.00
5 T	Bromomethane	0.403	0.383	5.0	96	-0.01
6 T	Chloroethane	0.419	0.413	1.4	95	0.00
7 T	Trichlorofluoromethane	0.959	0.946	1.4	93	0.00
8 T	Diethyl Ether	0.310	0.326	-5.2	102	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.452	0.428	5.3	90	0.00
10 T	Methyl Iodide	0.554	0.564	-1.8	96	0.00
11 T	Tert butyl alcohol	0.131	0.139	-6.1	103	0.00
12 CM	1,1-Dichloroethene	0.405	0.402	0.7#	93	0.00
13 T	Acrolein	0.021	0.019	9.5	99	0.00
14 T	Allyl chloride	0.896	0.899	-0.3	95	0.00
15 T	Acrylonitrile	0.335	0.361	-7.8	101	0.00
16 T	Acetone	0.391	0.348	11.0	84	0.00
17 T	Carbon Disulfide	0.958	0.911	4.9	90	0.00
18 T	Methyl Acetate	1.272	1.315	-3.4	101	0.00
19 T	Methyl tert-butyl Ether	1.715	1.769	-3.1	97	0.00
20 T	Methylene Chloride	0.555	0.523	5.8	95	0.00
21 T	trans-1,2-Dichloroethene	0.453	0.444	2.0	94	0.00
22 T	Diisopropyl ether	1.930	1.995	-3.4	97	0.00
23 T	Vinyl Acetate	1.493	1.400	6.2	85	0.00
24 P	1,1-Dichloroethane	0.983	0.970	1.3	94	0.00
25 T	2-Butanone	0.530	0.561	-5.8	98	0.00
26 T	2,2-Dichloropropane	0.769	0.735	4.4	87	0.00
27 T	cis-1,2-Dichloroethene	0.561	0.566	-0.9	95	0.00
28 T	Bromochloromethane	0.503	0.546	-8.5	107	0.00
29 T	Tetrahydrofuran	0.325	0.374	-15.1	104	0.00
30 C	Chloroform	1.008	1.048	-4.0#	98	0.00
31 T	Cyclohexane	0.990	0.927	6.4	94	0.00
32 T	1,1,1-Trichloroethane	0.841	0.887	-5.5	96	0.00
33 S	1,2-Dichloroethane-d4	0.725	0.734	-1.2	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	101	0.00
35 S	Dibromofluoromethane	0.300	0.296	1.3	102	0.00
36 T	1,1-Dichloropropene	0.421	0.416	1.2	94	0.00
37 T	Ethyl Acetate	0.569	0.606	-6.5	99	0.00
38 T	Carbon Tetrachloride	0.406	0.411	-1.2	96	0.00
39 T	Methylcyclohexane	0.523	0.511	2.3	91	0.00
40 TM	Benzene	1.299	1.300	-0.1	97	0.00
41 T	Methacrylonitrile	0.294	0.320	-8.8	110	0.00
42 TM	1,2-Dichloroethane	0.508	0.506	0.4	97	0.00
43 T	Isopropyl Acetate	0.913	0.980	-7.3	102	0.00
44 TM	Trichloroethene	0.311	0.308	1.0	96	0.00
45 C	1,2-Dichloropropane	0.347	0.355	-2.3#	98	0.00
46 T	Dibromomethane	0.232	0.236	-1.7	98	0.00
47 T	Bromodichloromethane	0.434	0.454	-4.6	96	0.00
48 T	Methyl methacrylate	0.430	0.441	-2.6	93	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110921\
 Data File : VN069430.D
 Acq On : 9 Nov 2021 20:36
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 10 16:20:36 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 09 18:18:22 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.009	-12.5	106	0.00
50 S	Toluene-d8	1.179	1.191	-1.0	101	0.00
51 T	4-Methyl-2-Pentanone	0.590	0.667	-13.1	103	0.00
52 CM	Toluene	0.809	0.833	-3.0#	97	0.00
53 T	t-1,3-Dichloropropene	0.453	0.482	-6.4	95	0.00
54 T	cis-1,3-Dichloropropene	0.495	0.520	-5.1	95	0.00
55 T	1,1,2-Trichloroethane	0.326	0.344	-5.5	99	0.00
56 T	Ethyl methacrylate	0.540	0.591	-9.4	100	0.00
57 T	1,3-Dichloropropane	0.573	0.590	-3.0	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.125	0.144	-15.2	106	0.00
59 T	2-Hexanone	0.439	0.495	-12.8	102	0.00
60 T	Dibromochloromethane	0.305	0.333	-9.2	97	0.00
61 T	1,2-Dibromoethane	0.330	0.356	-7.9	99	0.00
62 S	4-Bromofluorobenzene	0.433	0.436	-0.7	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	99	0.00
64 T	Tetrachloroethene	0.302	0.288	4.6	93	0.00
65 PM	Chlorobenzene	0.931	0.929	0.2	96	0.00
66 T	1,1,1,2-Tetrachloroethane	0.323	0.340	-5.3	97	0.00
67 C	Ethyl Benzene	1.716	1.728	-0.7#	96	0.00
68 T	m/p-Xylenes	0.631	0.639	-1.3	95	0.00
69 T	o-Xylene	0.629	0.648	-3.0	96	0.00
70 T	Styrene	1.011	1.062	-5.0	96	0.00
71 P	Bromoform	0.230	0.254	-10.4	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.981	3.918	1.6	95	0.00
74 T	N-amyl acetate	1.687	1.834	-8.7	100	0.00
75 P	1,1,2,2-Tetrachloroethane	1.343	1.362	-1.4	100	0.00
76 T	1,2,3-Trichloropropane	1.154	1.224	-6.1	102	0.00
77 T	Bromobenzene	0.910	0.911	-0.1	98	0.00
78 T	n-propylbenzene	4.480	4.448	0.7	94	0.00
79 T	2-Chlorotoluene	2.774	2.705	2.5	95	0.00
80 T	1,3,5-Trimethylbenzene	3.218	3.186	1.0	94	0.00
81 T	trans-1,4-Dichloro-2-butene	0.344	0.357	-3.8	97	0.00
82 T	4-Chlorotoluene	2.644	2.598	1.7	95	0.00
83 T	tert-Butylbenzene	2.820	2.831	-0.4	95	0.00
84 T	1,2,4-Trimethylbenzene	3.107	3.125	-0.6	94	0.00
85 T	sec-Butylbenzene	3.940	3.970	-0.8	94	0.00
86 T	p-Isopropyltoluene	3.128	3.148	-0.6	93	0.00
87 T	1,3-Dichlorobenzene	1.579	1.561	1.1	93	0.00
88 T	1,4-Dichlorobenzene	1.549	1.506	2.8	93	0.00
89 T	n-Butylbenzene	2.608	2.613	-0.2	91	0.00
90 T	Hexachloroethane	0.513	0.542	-5.7	95	0.00
91 T	1,2-Dichlorobenzene	1.502	1.534	-2.1	96	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.217	0.247	-13.8	100	0.00
93 T	1,2,4-Trichlorobenzene	0.662	0.727	-9.8	94	0.00
94 T	Hexachlorobutadiene	0.390	0.401	-2.8	93	0.00
95 T	Naphthalene	1.775	2.196	-23.7	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110921\
Data File : VN069430.D
Acq On : 9 Nov 2021 20:36
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 10 16:20:36 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110921W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 09 18:18:22 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.634	0.723	-14.0	98	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6