

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110922\
 Data File : VN075217.D
 Acq On : 09 Nov 2022 15:05
 Operator : JC\MD
 Sample : PB148908ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148908ZHE#03

Quant Time: Nov 10 02:43:13 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	103161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	197126	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	184583	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	76019	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	75115	54.041	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.080%	
35) Dibromofluoromethane	8.171	113	63094	54.522	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.040%	
50) Toluene-d8	10.571	98	239050	53.244	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 106.480%	
62) 4-Bromofluorobenzene	12.847	95	79926	50.677	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.360%	
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	12344	20.974	ug/l	94
18) Methyl Acetate	5.042	43	2138	1.112	ug/l #	74
20) Methylene Chloride	5.289	84	5627	4.288	ug/l #	82
37) Ethyl Acetate	7.571	43	18418	9.374	ug/l	98
43) Isopropyl Acetate	8.694	43	79066	25.943	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110922\
Data File : VN075217.D
Acq On : 09 Nov 2022 15:05
Operator : JC\MD
Sample : PB148908ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148908ZHE#03

Quant Time: Nov 10 02:43:13 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

