

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110922\
 Data File : VN075218.D
 Acq On : 09 Nov 2022 15:29
 Operator : JC\MD
 Sample : PB148908ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148908ZHE#04

Quant Time: Nov 10 02:43:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	100710	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	194354	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	183328	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	78111	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	76343	56.261	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 112.520%	
35) Dibromofluoromethane	8.171	113	62659	54.918	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.840%	
50) Toluene-d8	10.565	98	229689	51.888	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.780%	
62) 4-Bromofluorobenzene	12.847	95	79136	50.892	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.780%	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	11085	19.293	ug/l #	83
18) Methyl Acetate	5.047	43	2138	1.139	ug/l #	88
20) Methylene Chloride	5.283	84	5312	4.147	ug/l	92
37) Ethyl Acetate	7.565	43	17398	8.981	ug/l	99
43) Isopropyl Acetate	8.694	43	76571	25.483	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110922\
Data File : VN075218.D
Acq On : 09 Nov 2022 15:29
Operator : JC\MD
Sample : PB148908ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148908ZHE#04

Quant Time: Nov 10 02:43:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

