

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110922\
 Data File : VN075223.D
 Acq On : 09 Nov 2022 17:29
 Operator : JC\MD
 Sample : PB148908ZHE#09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148908ZHE#09

Quant Time: Nov 10 02:45:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	98837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	191694	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.864	117	181027	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	75833	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	74914	56.254	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 112.500%			
35) Dibromofluoromethane	8.171	113	60611	53.860	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 107.720%			
50) Toluene-d8	10.565	98	228457	52.326	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 104.660%			
62) 4-Bromofluorobenzene	12.847	95	79609	51.907	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 103.820%			
Target Compounds						Qvalue
16) Acetone	4.447	43	8468	15.017	ug/l	98
18) Methyl Acetate	5.035	43	2581	1.401	ug/l #	81
20) Methylene Chloride	5.277	84	4653	3.701	ug/l	88
37) Ethyl Acetate	7.571	43	17214	9.009	ug/l	99
43) Isopropyl Acetate	8.694	43	80887	27.293	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110922\
Data File : VN075223.D
Acq On : 09 Nov 2022 17:29
Operator : JC\MD
Sample : PB148908ZHE#09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148908ZHE#09

Quant Time: Nov 10 02:45:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

