

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110922\
 Data File : VN075224.D
 Acq On : 09 Nov 2022 17:53
 Operator : JC\MD
 Sample : PB148908ZHE#10
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148908ZHE#10

Quant Time: Nov 10 02:45:36 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	99216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	190906	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	180527	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	73267	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	73790	55.198	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.400%
35) Dibromofluoromethane	8.171	113	60354	53.853	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	107.700%
50) Toluene-d8	10.565	98	228893	52.642	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.280%
62) 4-Bromofluorobenzene	12.847	95	79779	52.232	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	104.460%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	8969	15.845	ug/l	96
18) Methyl Acetate	5.053	43	2593	1.402	ug/l #	75
20) Methylene Chloride	5.283	84	4932	3.908	ug/l #	87
37) Ethyl Acetate	7.565	43	17698	9.301	ug/l	99
43) Isopropyl Acetate	8.694	43	80863	27.397	ug/l #	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110922\
Data File : VN075224.D
Acq On : 09 Nov 2022 17:53
Operator : JC\MD
Sample : PB148908ZHE#10
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148908ZHE#10

Quant Time: Nov 10 02:45:36 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

