

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110922\
 Data File : VN075225.D
 Acq On : 09 Nov 2022 18:17
 Operator : JC\MD
 Sample : PB148908ZHE#11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148908ZHE#11

Quant Time: Nov 10 02:45:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	98500	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	186358	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	180070	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	73281	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	71679	54.009	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 108.020%	
35) Dibromofluoromethane	8.171	113	59962	54.809	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.620%	
50) Toluene-d8	10.565	98	223577	52.675	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 105.340%	
62) 4-Bromofluorobenzene	12.847	95	79004	52.987	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 105.980%	
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	11606	20.653	ug/l	96
18) Methyl Acetate	5.042	43	2932	1.597	ug/l #	68
20) Methylene Chloride	5.283	84	5687	4.539	ug/l	98
37) Ethyl Acetate	7.571	43	16861	9.077	ug/l	98
43) Isopropyl Acetate	8.694	43	85568	29.699	ug/l #	88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110922\
Data File : VN075225.D
Acq On : 09 Nov 2022 18:17
Operator : JC\MD
Sample : PB148908ZHE#11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148908ZHE#11

Quant Time: Nov 10 02:45:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

