

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
 Data File : VN080023.D
 Acq On : 09 Nov 2023 15:35
 Operator : JC\MD
 Sample : VN1109WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1109WBL01

Quant Time: Nov 10 00:30:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	339416	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	658135	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	662156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.800	152	229600	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	260053	50.290	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.580%
35) Dibromofluoromethane	8.171	113	232046	49.817	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.640%
50) Toluene-d8	10.571	98	811805	47.342	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.680%
62) 4-Bromofluorobenzene	12.853	95	278125	48.905	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.800%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
Data File : VN080023.D
Acq On : 09 Nov 2023 15:35
Operator : JC\MD
Sample : VN1109WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBL01

Quant Time: Nov 10 00:30:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

