

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
 Data File : VN080027.D
 Acq On : 09 Nov 2023 17:22
 Operator : JC\MD
 Sample : PB156960TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB156960TB

Quant Time: Nov 10 00:37:42 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	284162	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	582684	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	658935	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	245718	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	251975	58.202	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	116.400%	
35) Dibromofluoromethane	8.171	113	214675	52.055	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	104.120%	
50) Toluene-d8	10.570	98	788433	51.932	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	103.860%	
62) 4-Bromofluorobenzene	12.853	95	280047	55.619	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	111.240%	
Target Compounds						
16) Acetone	4.436	43	59095	38.860	ug/l	Qvalue 89
20) Methylene Chloride	5.283	84	13876	3.616	ug/l #	95
37) Ethyl Acetate	7.565	43	6314	1.169	ug/l #	91
43) Isopropyl Acetate	8.694	43	276686	29.215	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
Data File : VN080027.D
Acq On : 09 Nov 2023 17:22
Operator : JC\MD
Sample : PB156960TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB156960TB

Quant Time: Nov 10 00:37:42 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

