

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
 Data File : VN080031.D
 Acq On : 09 Nov 2023 18:58
 Operator : JC\MD
 Sample : 05284-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SB-1

Quant Time: Nov 10 00:40:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	300027	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	611666	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	672644	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	285581	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	254891	55.762	ug/l	0.00
Spiked Amount 50.000	Range	74 - 125	Recovery	= 111.520%		
35) Dibromofluoromethane	8.171	113	212899	49.178	ug/l	0.00
Spiked Amount 50.000	Range	75 - 124	Recovery	= 98.360%		
50) Toluene-d8	10.571	98	804086	50.454	ug/l	0.00
Spiked Amount 50.000	Range	86 - 113	Recovery	= 100.900%		
62) 4-Bromofluorobenzene	12.853	95	297295	56.247	ug/l	0.00
Spiked Amount 50.000	Range	83 - 123	Recovery	= 112.500%		
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	60261	37.531	ug/l	99
20) Methylene Chloride	5.277	84	13072	3.227	ug/l #	86
37) Ethyl Acetate	7.571	43	6934	1.223	ug/l #	87
40) Benzene	8.612	78	418445	22.602	ug/l	99
43) Isopropyl Acetate	8.694	43	284155	28.582	ug/l #	82
52) Toluene	10.635	92	259439	23.637	ug/l	98
67) Ethyl Benzene	11.965	91	28363	1.116	ug/l	95
68) m/p-Xylenes	12.065	106	41866	4.421	ug/l	89
69) o-Xylene	12.400	106	24689	2.707	ug/l	93
84) 1,2,4-Trimethylbenzene	13.488	105	22769	1.372	ug/l	79
95) Naphthalene	15.647	128	881629	58.673	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
Data File : VN080031.D
Acq On : 09 Nov 2023 18:58
Operator : JC\MD
Sample : 05284-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SB-1

Quant Time: Nov 10 00:40:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

