

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
 Data File : VN080033.D
 Acq On : 09 Nov 2023 19:46
 Operator : JC\MD
 Sample : 05295-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Nov 10 00:41:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	320779	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	645826	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	701744	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	279211	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	264637	54.149	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	108.300%
35) Dibromofluoromethane	8.171	113	212737	46.542	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.080%
50) Toluene-d8	10.571	98	838239	49.815	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.620%
62) 4-Bromofluorobenzene	12.853	95	305249	54.697	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	109.400%
Target Compounds						
						Qvalue
16) Acetone	4.436	43	65162	37.958	ug/l	94
20) Methylene Chloride	5.277	84	12547	2.897	ug/l	90
25) 2-Butanone	7.500	43	15153	6.028	ug/l	95
43) Isopropyl Acetate	8.694	43	281916	26.857	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
Data File : VN080033.D
Acq On : 09 Nov 2023 19:46
Operator : JC\MD
Sample : 05295-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Nov 10 00:41:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

