

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
 Data File : VN080034.D
 Acq On : 09 Nov 2023 20:10
 Operator : JC\MD
 Sample : 05311-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-1

Quant Time: Nov 10 00:41:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	316141	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	676247	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	720633	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	278127	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	270462	56.153	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.300%
35) Dibromofluoromethane	8.171	113	228128	47.664	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.320%
50) Toluene-d8	10.571	98	869815	49.366	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.740%
62) 4-Bromofluorobenzene	12.853	95	302686	51.798	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.600%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	59421	35.122	ug/l	90
20) Methylene Chloride	5.277	84	17225	4.035	ug/l	89
37) Ethyl Acetate	7.571	43	7184	1.146	ug/l #	75
43) Isopropyl Acetate	8.694	43	299806	27.277	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
Data File : VN080034.D
Acq On : 09 Nov 2023 20:10
Operator : JC\MD
Sample : 05311-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-1

Quant Time: Nov 10 00:41:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

