

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
 Data File : VN080035.D
 Acq On : 09 Nov 2023 20:34
 Operator : JC\MD
 Sample : 05311-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-2

Quant Time: Nov 10 00:42:00 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	300456	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	640565	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	692176	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	270191	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	260158	56.833	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	113.660%
35) Dibromofluoromethane	8.171	113	220983	48.743	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.480%
50) Toluene-d8	10.571	98	829246	49.685	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.380%
62) 4-Bromofluorobenzene	12.853	95	293339	52.995	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.980%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	58968	36.674	ug/l	95
18) Methyl Acetate	5.036	43	9210	1.197	ug/l #	77
20) Methylene Chloride	5.283	84	15811	3.897	ug/l	93
37) Ethyl Acetate	7.559	43	7591	1.279	ug/l #	86
43) Isopropyl Acetate	8.694	43	307346	29.520	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
Data File : VN080035.D
Acq On : 09 Nov 2023 20:34
Operator : JC\MD
Sample : 05311-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-2

Quant Time: Nov 10 00:42:00 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

