

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
 Data File : VN080036.D
 Acq On : 09 Nov 2023 20:58
 Operator : JC\MD
 Sample : 05311-09
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-3

Quant Time: Nov 10 00:42:25 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	298838	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	637244	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	684768	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	263432	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	260203	57.151	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.300%
35) Dibromofluoromethane	8.171	113	217038	48.122	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.240%
50) Toluene-d8	10.571	98	826098	49.754	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.500%
62) 4-Bromofluorobenzene	12.853	95	289419	52.559	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.120%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	39967	24.991	ug/l	97
20) Methylene Chloride	5.283	84	13960	3.459	ug/l	85
37) Ethyl Acetate	7.571	43	6415	1.086	ug/l #	81
43) Isopropyl Acetate	8.694	43	323888	31.271	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
Data File : VN080036.D
Acq On : 09 Nov 2023 20:58
Operator : JC\MD
Sample : 05311-09
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-3

Quant Time: Nov 10 00:42:25 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

