

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
 Data File : VN080037.D
 Acq On : 09 Nov 2023 21:22
 Operator : JC\MD
 Sample : 05311-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4

Quant Time: Nov 10 00:42:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	294589	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	626012	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	664734	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	243211	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	254565	56.719	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	113.440%	
35) Dibromofluoromethane	8.171	113	215703	48.684	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	97.360%	
50) Toluene-d8	10.570	98	800182	49.058	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.120%	
62) 4-Bromofluorobenzene	12.853	95	268860	49.701	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	=	99.400%	
Target Compounds						
					Qvalue	
16) Acetone	4.441	43	60543	38.403	ug/l	91
18) Methyl Acetate	5.041	43	9852	1.306	ug/l #	79
20) Methylene Chloride	5.283	84	16552	4.161	ug/l #	87
43) Isopropyl Acetate	8.694	43	286710	28.178	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
Data File : VN080037.D
Acq On : 09 Nov 2023 21:22
Operator : JC\MD
Sample : 05311-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4

Quant Time: Nov 10 00:42:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

