

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
 Data File : VN080038.D
 Acq On : 09 Nov 2023 21:45
 Operator : JC\MD
 Sample : 05311-15
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-5

Manual Integrations
 APPROVED

Reviewed By : John Carlone 11/10/2023
 Supervised By : Mahesh Dadoda 11/10/2023

Quant Time: Nov 10 00:43:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	282945	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	609057	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	670151	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	254624	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	256687	59.546	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 119.100%			
35) Dibromofluoromethane	8.171	113	215488	49.990	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.980%			
50) Toluene-d8	10.571	98	801897	50.532	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.060%			
62) 4-Bromofluorobenzene	12.853	95	282649	53.705	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 107.420%			
Target Compounds					Qvalue	
16) Acetone	4.430	43	56361	37.222	ug/l	97
18) Methyl Acetate	5.030	43	10005m	1.381	ug/l	
20) Methylene Chloride	5.277	84	16579	4.339	ug/l	92
37) Ethyl Acetate	7.571	43	7454	1.321	ug/l #	86
43) Isopropyl Acetate	8.694	43	310816	31.398	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
Data File : VN080038.D
Acq On : 09 Nov 2023 21:45
Operator : JC\MD
Sample : 05311-15
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-5

Quant Time: Nov 10 00:43:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 11/10/2023
Supervised By :Mahesh Dadoda 11/10/2023

