

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
 Data File : VN080039.D
 Acq On : 09 Nov 2023 22:09
 Operator : JC\MD
 Sample : 05311-18
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-6

Quant Time: Nov 10 00:43:38 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	285744	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	617197	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	668532	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	250697	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	253375	58.202	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	116.400%
35) Dibromofluoromethane	8.171	113	214163	49.027	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.060%
50) Toluene-d8	10.571	98	803066	49.938	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.880%
62) 4-Bromofluorobenzene	12.853	95	281676	52.814	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	105.620%
Target Compounds						
					Qvalue	
16) Acetone	4.436	43	58311	38.132	ug/l #	82
18) Methyl Acetate	5.030	43	9291	1.270	ug/l #	61
20) Methylene Chloride	5.271	84	15703	4.070	ug/l #	85
43) Isopropyl Acetate	8.694	43	296517	29.558	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
Data File : VN080039.D
Acq On : 09 Nov 2023 22:09
Operator : JC\MD
Sample : 05311-18
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-6

Quant Time: Nov 10 00:43:38 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

