

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
 Data File : VN080026.D
 Acq On : 09 Nov 2023 16:59
 Operator : JC\MD
 Sample : VN1109WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1109WBSD01

Manual Integrations
 APPROVED

Reviewed By :John
 Carlone

Quant Time: Nov 10 00:36:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

11/10/2023
 Supervised By :Mahesh
 Dadoda

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	300567	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	511491	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	468419	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	214591	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	252220	55.079	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.160%	
35) Dibromofluoromethane	8.171	113	198361	54.794	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 109.580%	
50) Toluene-d8	10.571	98	692822	51.986	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.980%	
62) 4-Bromofluorobenzene	12.853	95	245089	55.451	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 110.900%	

Target Compounds					Qvalue	
2) Dichlorodifluoromethane	2.124	85	62498	17.263	ug/l	97
3) Chloromethane	2.359	50	71851	15.087	ug/l	99
4) Vinyl Chloride	2.512	62	107468	22.677	ug/l	99
5) Bromomethane	2.959	94	36016	12.508	ug/l	94
6) Chloroethane	3.118	64	54070	14.439	ug/l	97
7) Trichlorofluoromethane	3.495	101	111965	18.995	ug/l	96
8) Diethyl Ether	3.965	74	42425	18.967	ug/l	65
9) 1,1,2-Trichlorotrifluo...	4.371	101	66734	19.530	ug/l	99
10) Methyl Iodide	4.589	142	64220	16.458	ug/l	98
11) Tert butyl alcohol	5.524	59	56239	96.208	ug/l	96
12) 1,1-Dichloroethene	4.342	96	57201	17.680	ug/l #	80
13) Acrolein	4.183	56	55030	108.505	ug/l	96
14) Allyl chloride	5.018	41	107656	21.726	ug/l #	86
15) Acrylonitrile	5.718	53	197740	108.548	ug/l	98
16) Acetone	4.430	43	179237	111.431	ug/l	95
17) Carbon Disulfide	4.712	76	148078	15.868	ug/l	99
18) Methyl Acetate	5.024	43	170019	22.093	ug/l #	84
19) Methyl tert-butyl Ether	5.800	73	227224	20.400	ug/l	94
20) Methylene Chloride	5.283	84	79905	19.687	ug/l #	87
21) trans-1,2-Dichloroethene	5.789	96	66427	18.410	ug/l #	80
22) Diisopropyl ether	6.677	45	249235	21.704	ug/l #	89
23) Vinyl Acetate	6.606	43	627858	99.122	ug/l #	87
24) 1,1-Dichloroethane	6.565	63	144104	20.646	ug/l	96
25) 2-Butanone	7.483	43	255109	108.307	ug/l #	82
26) 2,2-Dichloropropane	7.489	77	113141	19.600	ug/l	98
27) cis-1,2-Dichloroethene	7.494	96	82753	19.173	ug/l	87
28) Bromochloromethane	7.818	49	84263	24.773	ug/l #	75
29) Tetrahydrofuran	7.847	42	165984	108.312	ug/l #	76
30) Chloroform	7.971	83	155063	20.453	ug/l	97
31) Cyclohexane	8.259	56	105771	17.480	ug/l	85
32) 1,1,1-Trichloroethane	8.171	97	126227	20.434	ug/l	95
36) 1,1-Dichloropropene	8.371	75	101775	19.690	ug/l	95
37) Ethyl Acetate	7.559	43	109380	23.073	ug/l #	90
38) Carbon Tetrachloride	8.365	117	108505	20.829	ug/l	93
39) Methylcyclohexane	9.606	83	84890	17.451	ug/l #	84
40) Benzene	8.612	78	316923	20.471	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
 Data File : VN080026.D
 Acq On : 09 Nov 2023 16:59
 Operator : JC\MD
 Sample : VN1109WBSD01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1109WBSD01

Manual Integrations
 APPROVED

Reviewed By :John
 Carlone

Quant Time: Nov 10 00:36:47 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 02 05:49:08 2023
 Response via : Initial Calibration

11/10/2023
 Supervised By :Mahesh
 Dadoda

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	55687	22.121	ug/l #	85
42) 1,2-Dichloroethane	8.677	62	122815	22.711	ug/l	98
43) Isopropyl Acetate	8.694	43	166842	20.069	ug/l #	92
44) Trichloroethene	9.359	130	76982	20.039	ug/l	93
45) 1,2-Dichloropropane	9.624	63	87937	22.226	ug/l	95
46) Dibromomethane	9.712	93	58362	21.603	ug/l	96
47) Bromodichloromethane	9.894	83	122605	21.280	ug/l	99
48) Methyl methacrylate	9.688	41	80827	22.197	ug/l #	79
49) 1,4-Dioxane	9.700	88	29238	454.407	ug/l #	78
51) 4-Methyl-2-Pentanone	10.447	43	544785	116.397	ug/l	90
52) Toluene	10.635	92	191259	20.838	ug/l	98
53) t-1,3-Dichloropropene	10.841	75	117609	20.823	ug/l	99
54) cis-1,3-Dichloropropene	10.312	75	131077	21.292	ug/l #	89
55) 1,1,2-Trichloroethane	11.018	97	82620	22.380	ug/l	96
56) Ethyl methacrylate	10.877	69	110735	21.029	ug/l #	86
57) 1,3-Dichloropropane	11.165	76	144002	22.130	ug/l	100
58) 2-Chloroethyl Vinyl ether	10.165	63	244413	91.579	ug/l	91
59) 2-Hexanone	11.200	43	391561	109.359	ug/l	88
60) Dibromochloromethane	11.365	129	85476	21.631	ug/l	99
61) 1,2-Dibromoethane	11.471	107	80259	21.661	ug/l	96
64) Tetrachloroethene	11.106	164	73344	24.176	ug/l	95
65) Chlorobenzene	11.894	112	201252	19.806	ug/l	96
66) 1,1,1,2-Tetrachloroethane	11.965	131	79307	21.663	ug/l	97
67) Ethyl Benzene	11.965	91	339198	19.161	ug/l	95
68) m/p-Xylenes	12.076	106	259801	39.396	ug/l	93
69) o-Xylene	12.406	106	125229	19.720	ug/l	94
70) Styrene	12.418	104	206295	20.214	ug/l	96
71) Bromoform	12.582	173	58791	22.616	ug/l #	96
73) Isopropylbenzene	12.700	105	306902	18.844	ug/l	100
74) N-amyl acetate	12.500	43	125864	17.706	ug/l	90
75) 1,1,2,2-Tetrachloroethane	12.941	83	120695	20.902	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	111179m	22.068	ug/l	
77) Bromobenzene	12.988	156	81523	20.446	ug/l	92
78) n-propylbenzene	13.041	91	366798	19.667	ug/l	98
79) 2-Chlorotoluene	13.129	91	240962	20.225	ug/l	100
80) 1,3,5-Trimethylbenzene	13.176	105	257051	20.286	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.741	75	33180	17.919	ug/l #	74
82) 4-Chlorotoluene	13.229	91	242165	20.454	ug/l	98
83) tert-Butylbenzene	13.441	119	209290	20.178	ug/l	99
84) 1,2,4-Trimethylbenzene	13.488	105	258119	20.705	ug/l	99
85) sec-Butylbenzene	13.623	105	282849	19.683	ug/l	99
86) p-Isopropyltoluene	13.735	119	225756	19.269	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	141788	19.115	ug/l	99
88) 1,4-Dichlorobenzene	13.818	146	145505	19.326	ug/l	99
89) n-Butylbenzene	14.059	91	189958	18.825	ug/l	97
90) Hexachloroethane	14.335	117	45486	19.604	ug/l	93
91) 1,2-Dichlorobenzene	14.112	146	141629	20.227	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.723	75	21132	21.580	ug/l	89
93) 1,2,4-Trichlorobenzene	15.400	180	62424	18.749	ug/l	98
94) Hexachlorobutadiene	15.506	225	29943	19.839	ug/l	97
95) Naphthalene	15.647	128	180147	15.955	ug/l	98
96) 1,2,3-Trichlorobenzene	15.847	180	61451	18.829	ug/l	98

11/10/2023

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN110923\
Data File : VN080026.D
Acq On : 09 Nov 2023 16:59
Operator : JC\MD
Sample : VN1109WBSD01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Quant Time: Nov 10 00:36:47 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N110123W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 02 05:49:08 2023
Response via : Initial Calibration

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01

Manual Integrations
APPROVED


Reviewed By :John
Carlone

11/10/2023
Supervised By :Mahesh
Dadoda

11/10/2023

Compound	R.T.	QI	on	Response	Conc	Units	Dev(Min)
----------	------	----	----	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 1067
Delta R.T. 0.001 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion:168 Resp: 30056
Ion Ratio Lower Upper
168 100
99 62.0 53.0 79.6

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda

11/10/2023



#2
Dichlorodifluoromethane
Concen: 17.263 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion: 85 Resp: 62498
Ion Ratio Lower Upper
85 100
87 30.2 15.8 47.4





#3
Chloromethane
Concen: 15.087 ug/l
RT: 2.359 min Scan# 69
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion: 50 Resp: 7185
Ion Ratio Lower Upper
50 100
52 33.4 26.5 39.7

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda

11/10/2023



#4
Vinyl Chloride
Concen: 22.677 ug/l
RT: 2.512 min Scan# 95
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion: 62 Resp: 107468
Ion Ratio Lower Upper
62 100
64 32.0 26.2 39.2





#5
Bromomethane
Concen: 12.508 ug/l
RT: 2.959 min Scan# 171
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion: 94 Resp: 36010
Ion Ratio Lower Upper
94 100
96 85.4 73.1 109.7

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#6
Chloroethane
Concen: 14.439 ug/l
RT: 3.118 min Scan# 198
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion: 64 Resp: 54070
Ion Ratio Lower Upper
64 100
66 33.3 25.4 38.2





#7
Trichlorofluoromethane
Concen: 18.995 ug/l
RT: 3.495 min Scan# 262
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion:101 Resp: 11196
Ion Ratio Lower Upper
101 100
103 66.9 50.7 76.1

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#8
Diethyl Ether
Concen: 18.967 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion: 74 Resp: 42425
Ion Ratio Lower Upper
74 100
45 97.3 34.4 103.3





#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.530 ug/l
RT: 4.371 min Scan# 411
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion:101 Resp: 66734
Ion Ratio Lower Upper
101 100
85 44.9 36.6 55.0
151 71.4 57.6 86.4

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#10
Methyl Iodide
Concen: 16.458 ug/l
RT: 4.589 min Scan# 448
Delta R.T. -0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion:142 Resp: 64220
Ion Ratio Lower Upper
142 100
127 45.1 37.6 56.4
141 14.6 11.4 17.0





#11
Tert butyl alcohol
Concen: 96.208 ug/l
RT: 5.524 min Scan# 607
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion: 59 Resp: 56239
Ion Ratio Lower Upper
59 100
57 11.9 8.2 12.4

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#12
1,1-Dichloroethene
Concen: 17.680 ug/l
RT: 4.342 min Scan# 406
Delta R.T. -0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion: 96 Resp: 57201
Ion Ratio Lower Upper
96 100
61 171.9 112.7 169.1#
98 71.5 51.2 76.8





#13

Acrolein

Concen: 108.505 ug/l

RT: 4.183 min Scan# 31

Delta R.T. -0.000 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59

Instrument :

MSVOA_N

ClientSampleId :

VN1109WBSD01

Tgt Ion: 56 Resp: 55030

Ion Ratio Lower Upper

56 100

55 71.0 54.5 81.7

Manual Integrations





APPROVED

Reviewed By :John
Carlone

11/10/2023

Supervised By :Mahesh
Dadoda

11/10/2023



#14

Allyl chloride

Concen: 21.726 ug/l

RT: 5.018 min Scan# 521

Delta R.T. -0.006 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59




Tgt Ion: 41 Resp: 107656


Ion Ratio Lower Upper

41 100

39 66.0 58.8 88.2


76 33.3 39.1 58.7#





#15
Acrylonitrile
Concen: 108.548 ug/l
RT: 5.718 min Scan# 641
Delta R.T. -0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion: 53 Resp: 197740
Ion Ratio Lower Upper
53 100
52 83.0 65.3 97.9
51 35.2 28.9 43.3

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#16
Acetone
Concen: 111.431 ug/l
RT: 4.430 min Scan# 421
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion: 43 Resp: 179237
Ion Ratio Lower Upper
43 100
58 31.8 28.0 42.0





#17
Carbon Disulfide
Concen: 15.868 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion: 76 Resp: 148078
Ion Ratio Lower Upper
76 100
78 9.7 7.4 11.0

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#18
Methyl Acetate
Concen: 22.093 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion: 43 Resp: 170019
Ion Ratio Lower Upper
43 100
74 23.5 26.2 39.4#





#19
Methyl tert-butyl Ether
Concen: 20.400 ug/l
RT: 5.800 min Scan# 61
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion: 73 Resp: 227224
Ion Ratio Lower Upper
73 100
57 22.4 15.6 23.4

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#20
Methylene Chloride
Concen: 19.687 ug/l
RT: 5.283 min Scan# 566
Delta R.T. 0.012 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion: 84 Resp: 79905
Ion Ratio Lower Upper
84 100
49 120.0 79.7 119.5#
51 38.1 25.2 37.8#
86 65.1 53.3 79.9





#21
trans-1,2-Dichloroethene
Concen: 18.410 ug/l
RT: 5.789 min Scan# 611
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion: 96 Resp: 6642
Ion Ratio Lower Upper
96 100
61 141.5 90.1 135.1
98 62.5 46.4 69.6

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#22
Diisopropyl ether
Concen: 21.704 ug/l
RT: 6.677 min Scan# 803
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion: 45 Resp: 249235
Ion Ratio Lower Upper
45 100
43 54.2 45.5 68.3
87 26.9 31.8 47.8#
59 10.9 11.6 17.4#





#23

Vinyl Acetate

Concen: 99.122 ug/l

RT: 6.606 min Scan# 791

Delta R.T. 0.000 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59

Instrument :

MSVOA_N

ClientSampleId :

VN1109WBSD01

Tgt Ion: 43 Resp: 62785

Ion Ratio Lower Upper

43 100

86 10.7 13.0 19.6

Manual Integrations





APPROVED

Reviewed By :John
Carlone

11/10/2023

Supervised By :Mahesh
Dadoda

11/10/2023



#24

1,1-Dichloroethane

Concen: 20.646 ug/l

RT: 6.565 min Scan# 784

Delta R.T. -0.006 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59




Tgt Ion: 63 Resp: 144104


Ion Ratio Lower Upper

63 100

98 6.2 3.9 11.7


100 3.9 2.5 7.4





#25
2-Butanone
Concen: 108.307 ug/l
RT: 7.483 min Scan# 940
Delta R.T. -0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion: 43 Resp: 255109
Ion Ratio Lower Upper
43 100
72 26.6 30.2 45.2

Manual Integrations
APPROVED


Reviewed By :John
Carlone

11/10/2023
Supervised By :Mahesh
Dadoda



11/10/2023




#26
2,2-Dichloropropane
Concen: 19.600 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59




Tgt Ion: 77 Resp: 113141
Ion Ratio Lower Upper
77 100
97 21.4 11.2 33.6





#27
 cis-1,2-Dichloroethene
 Concen: 19.173 ug/l
 RT: 7.494 min Scan# 941
 Delta R.T. 0.006 min
 Lab File: VN080026.D
 Acq: 09 Nov 2023 16:59

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1109WBSD01






Tgt Ion: 96 Resp: 8275
 Ion Ratio Lower Upper
 96 100
 61 150.2 0.0 255.2
 98 64.2 0.0 128.8

Manual Integrations
 APPROVED

Reviewed By :John
 Carlone




11/10/2023
 Supervised By :Mahesh
 Dadoda


11/10/2023



#28
 Bromochloromethane
 Concen: 24.773 ug/l
 RT: 7.818 min Scan# 997
 Delta R.T. 0.006 min
 Lab File: VN080026.D
 Acq: 09 Nov 2023 16:59

Tgt Ion: 49 Resp: 84263
 Ion Ratio Lower Upper
 49 100
 129 1.8 0.0 4.8
 130 69.6 75.3 112.9#





#29

Tetrahydrofuran

Concen: 108.312 ug/l

RT: 7.847 min Scan# 1001

Delta R.T. 0.006 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59

Instrument :

MSVOA_N

ClientSampleId :

VN1109WBSD01

Tgt Ion: 42 Resp: 165984

Ion	Ratio	Lower	Upper
42	100		
72	43.4	49.8	74.6
71	40.9	47.4	71.2





Manual Integrations
APPROVED

Reviewed By :John
Carlone

11/10/2023

Supervised By :Mahesh
Dadoda

11/10/2023



#30

Chloroform

Concen: 20.453 ug/l

RT: 7.971 min Scan# 1023




Delta R.T. 0.000 min


Lab File: VN080026.D

Acq: 09 Nov 2023 16:59

Tgt Ion: 83 Resp: 155063


Ion	Ratio	Lower	Upper
83	100		
85	66.5	55.4	83.0





#31
Cyclohexane
Concen: 17.480 ug/l
RT: 8.259 min Scan# 1072
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
Client Sample Id :
VN1109WBSD01






Tgt Ion: 56 Resp: 105772
Ion Ratio Lower Upper
56 100
69 32.0 26.8 40.2
84 82.9 81.8 122.6

Manual Integrations
APPROVED


Reviewed By :John
Carlone

11/10/2023
Supervised By :Mahesh
Dadoda



11/10/2023




#32
1,1,1-Trichloroethane
Concen: 20.434 ug/l
RT: 8.171 min Scan# 1057
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59



Tgt Ion: 97 Resp: 126227
Ion Ratio Lower Upper
97 100
99 61.8 50.3 75.5
61 49.2 34.7 52.1





#33

1,2-Dichloroethane-d4

Concen: 55.079 ug/l

RT: 8.583 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59

Instrument :

MSVOA_N

ClientSampleId :

VN1109WBSD01

Tgt Ion: 65 Resp: 252220

Ion Ratio Lower Upper

65 100

67 52.0 0.0 106.6

Manual Integrations





APPROVED

Reviewed By :John
Carlone

11/10/2023

Supervised By :Mahesh
Dadoda

11/10/2023



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 9.106 min Scan# 1216

Delta R.T. 0.000 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59




Tgt Ion:114 Resp: 511491


Ion Ratio Lower Upper

114 100

63 21.3 0.0 36.4

88 15.4 0.0 30.6





#35

Dibromofluoromethane

Concen: 54.794 ug/l

RT: 8.171 min Scan# 1056

Delta R.T. 0.000 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59

Instrument :

MSVOA_N

ClientSampleId :

VN1109WBSD01

Tgt Ion:113 Resp: 19836

Ion Ratio Lower Upper

113 100

111 101.1 82.6 123.8

192 17.0 12.7 19.1

Manual Integrations





APPROVED

Reviewed By :John
Carlone

11/10/2023

Supervised By :Mahesh
Dadoda

11/10/2023



#36

1,1-Dichloropropene

Concen: 19.690 ug/l

RT: 8.371 min Scan# 1091

Delta R.T. -0.006 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59




Tgt Ion: 75 Resp: 101775


Ion Ratio Lower Upper

75 100

110 33.1 19.1 57.3


77 31.4 25.0 37.4





#37
Ethyl Acetate
Concen: 23.073 ug/l
RT: 7.559 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion: 43 Resp: 109380
Ion Ratio Lower Upper
43 100
61 12.0 13.2 19.8
70 9.9 11.3 16.9

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#38
Carbon Tetrachloride
Concen: 20.829 ug/l
RT: 8.365 min Scan# 1090
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion:117 Resp: 108505
Ion Ratio Lower Upper
117 100
119 87.6 77.3 115.9
121 28.5 23.3 34.9





#39

Methylcyclohexane

Concen: 17.451 ug/l

RT: 9.606 min Scan# 1130

Delta R.T. 0.000 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59

Instrument :

MSVOA_N

ClientSampleId :

VN1109WBSD01

Tgt Ion: 83 Resp: 84890

Ion	Ratio	Lower	Upper
83	100		
55	81.9	49.1	73.7
98	48.7	40.6	61.0





Manual Integrations
APPROVED

Reviewed By :John
Carlone

11/10/2023

Supervised By :Mahesh
Dadoda

11/10/2023



#40

Benzene

Concen: 20.471 ug/l

RT: 8.612 min Scan# 1132




Delta R.T. 0.000 min


Lab File: VN080026.D

Acq: 09 Nov 2023 16:59

Tgt Ion: 78 Resp: 316923


Ion	Ratio	Lower	Upper
78	100		
77	23.6	19.0	28.6





#41
Methacrylonitrile
Concen: 22.121 ug/l
RT: 7.783 min Scan# 991
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion: 41 Resp: 5568
Ion Ratio Lower Upper
41 100
39 58.5 42.6 64.0
67 75.0 77.5 116.3
52 32.8 28.9 43.3

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#42
1,2-Dichloroethane
Concen: 22.711 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion: 62 Resp: 122815
Ion Ratio Lower Upper
62 100
98 9.0 0.0 19.8





#43
Isopropyl Acetate
Concen: 20.069 ug/l
RT: 8.694 min Scan# 1145
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion: 43 Resp: 16684
Ion Ratio Lower Upper
43 100
61 27.6 24.7 37.1
87 13.1 14.6 21.8

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#44
Trichloroethene
Concen: 20.039 ug/l
RT: 9.359 min Scan# 1259
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion:130 Resp: 76982
Ion Ratio Lower Upper
130 100
95 95.1 0.0 204.6





#45

1,2-Dichloropropane

Concen: 22.226 ug/l

RT: 9.624 min Scan# 1304

Delta R.T. 0.000 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59

Instrument :

MSVOA_N

ClientSampleId :

VN1109WBSD01

Tgt Ion: 63 Resp: 8793

Ion Ratio Lower Upper

63 100

65 27.2 23.9 35.9

Manual Integrations





APPROVED

Reviewed By :John
Carlone

11/10/2023

Supervised By :Mahesh
Dadoda

11/10/2023



#46

Dibromomethane

Concen: 21.603 ug/l

RT: 9.712 min Scan# 1319

Delta R.T. 0.000 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59




Tgt Ion: 93 Resp: 58362


Ion Ratio Lower Upper

93 100

95 81.8 63.8 95.8

174 87.9 65.6 98.4





#47

Bromodichloromethane

Concen: 21.280 ug/l

RT: 9.894 min Scan# 1350

Delta R.T. 0.006 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59

Instrument :

MSVOA_N

ClientSampleId :

VN1109WBSD01

Tgt Ion: 83 Resp: 12260

Ion	Ratio	Lower	Upper
83	100		
85	63.0	49.6	74.4
127	8.1	7.0	10.6





Manual Integrations
APPROVED

Reviewed By :John
Carlone

11/10/2023

Supervised By :Mahesh
Dadoda

11/10/2023



#48

Methyl methacrylate

Concen: 22.197 ug/l

RT: 9.688 min Scan# 1315




Delta R.T. 0.006 min


Lab File: VN080026.D

Acq: 09 Nov 2023 16:59

Tgt Ion: 41 Resp: 80827

Ion	Ratio	Lower	Upper
41	100		
69	85.8	90.4	135.6
39	57.1	54.6	81.8





#49

1,4-Dioxane

Concen: 454.407 ug/l

RT: 9.700 min Scan# 11

Delta R.T. 0.006 min

Lab File: VN080026.D

Acq: 09 Nov 2023 16:59

Instrument :

MSVOA_N

ClientSampleId :

VN1109WBSD01

Tgt Ion: 88 Resp: 29238





Ion	Ratio	Lower	Upper
88	100		
43	32.8	17.0	25.6
58	74.4	47.0	70.4

Manual Integrations
APPROVED

Reviewed By :John
Carlone

11/10/2023
Supervised By :Mahesh
Dadoda

11/10/2023



#50

Toluene-d8

Concen: 51.986 ug/l

RT: 10.571 min Scan# 1465




Delta R.T. 0.000 min


Lab File: VN080026.D

Acq: 09 Nov 2023 16:59

Tgt Ion: 98 Resp: 692822

Ion	Ratio	Lower	Upper
98	100		
100	64.5	52.0	78.0





#51
4-Methyl-2-Pentanone
Concen: 116.397 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion: 43 Resp: 544785
Ion Ratio Lower Upper
43 100
58 38.8 36.6 54.8

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#52
Toluene
Concen: 20.838 ug/l
RT: 10.635 min Scan# 1476
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion: 92 Resp: 191259
Ion Ratio Lower Upper
92 100
91 175.1 137.4 206.0





#53
t-1,3-Dichloropropene
Concen: 20.823 ug/l
RT: 10.841 min Scan# 1510
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion: 75 Resp: 117609
Ion Ratio Lower Upper
75 100
77 31.1 25.4 38.0

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#54
cis-1,3-Dichloropropene
Concen: 21.292 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion: 75 Resp: 131077
Ion Ratio Lower Upper
75 100
77 28.4 26.1 39.1
39 47.6 31.6 47.4#





#55
1,1,2-Trichloroethane
Concen: 22.380 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion	Ratio	Lower	Upper
97	100		
83	87.6	65.1	97.7
85	54.3	43.4	65.2
99	58.2	49.2	73.8

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda

11/10/2023



#56
Ethyl methacrylate
Concen: 21.029 ug/l
RT: 10.877 min Scan# 1517
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion	Ratio	Lower	Upper
69	100		
41	65.4	39.6	59.4
39	36.1	28.6	43.0





#57
1,3-Dichloropropane
Concen: 22.130 ug/l
RT: 11.165 min Scan# 1566
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion: 76 Resp: 14400
Ion Ratio Lower Upper
76 100
78 32.6 26.1 39.1

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#58
2-Chloroethyl Vinyl ether
Concen: 91.579 ug/l
RT: 10.165 min Scan# 1396
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion: 63 Resp: 244413
Ion Ratio Lower Upper
63 100
106 27.6 26.2 39.4





#59
2-Hexanone
Concen: 109.359 ug/l
RT: 11.200 min Scan# 1110
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion: 43 Resp: 39156
Ion Ratio Lower Upper
43 100
58 53.7 31.7 95.1

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#60
Dibromochloromethane
Concen: 21.631 ug/l
RT: 11.365 min Scan# 1600
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion:129 Resp: 85476
Ion Ratio Lower Upper
129 100
127 76.2 38.6 116.0





#61
1,2-Dibromoethane
Concen: 21.661 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion:107 Resp: 80259
Ion Ratio Lower Upper
107 100
109 92.0 76.6 114.8

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#62
4-Bromofluorobenzene
Concen: 55.451 ug/l
RT: 12.853 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion: 95 Resp: 245089
Ion Ratio Lower Upper
95 100
174 71.2 0.0 143.8
176 68.0 0.0 135.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 1686
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion:117 Resp: 468419
Ion Ratio Lower Upper
117 100
82 59.4 44.0 66.0
119 33.3 25.6 38.4

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#64
Tetrachloroethene
Concen: 24.176 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion:164 Resp: 73344
Ion Ratio Lower Upper
164 100
166 134.7 101.4 152.0
129 108.8 82.4 123.6
131 100.4 78.2 117.4





#65
Chlorobenzene
Concen: 19.806 ug/l
RT: 11.894 min Scan# 1112
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion:112 Resp: 201252
Ion Ratio Lower Upper
112 100
114 33.7 25.0 37.4

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#66
1,1,1,2-Tetrachloroethane
Concen: 21.663 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion:131 Resp: 79307
Ion Ratio Lower Upper
131 100
133 96.1 46.4 139.1
119 65.7 34.1 102.2





#67
Ethyl Benzene
Concen: 19.161 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion: 91 Resp: 339198
Ion Ratio Lower Upper
91 100
106 29.0 25.6 38.4

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#68
m/p-Xylenes
Concen: 39.396 ug/l
RT: 12.076 min Scan# 1721
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion:106 Resp: 259801
Ion Ratio Lower Upper
106 100
91 209.7 159.6 239.4





#69
o-Xylene
Concen: 19.720 ug/l
RT: 12.406 min Scan# 1776
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion:106 Resp: 125229
Ion Ratio Lower Upper
106 100
91 219.3 105.2 315.5

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#70
Styrene
Concen: 20.214 ug/l
RT: 12.418 min Scan# 1779
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion:104 Resp: 206295
Ion Ratio Lower Upper
104 100
78 53.4 39.5 59.3
103 56.1 43.2 64.8





#71
Bromoform
Concen: 22.616 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion:173 Resp: 5879
Ion Ratio Lower Upper
173 100
175 46.0 24.3 73.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 1313
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion:152 Resp: 214591
Ion Ratio Lower Upper
152 100
115 62.8 32.3 96.8
150 163.8 0.0 355.8





#73
Isopropylbenzene
Concen: 18.844 ug/l
RT: 12.700 min Scan# 1827
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion:105 Resp: 306902
Ion Ratio Lower Upper
105 100
120 26.0 13.1 39.1

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#74
N-ethyl acetate
Concen: 17.706 ug/l
RT: 12.500 min Scan# 1793
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion: 43 Resp: 125864
Ion Ratio Lower Upper
43 100
70 44.5 43.4 65.2
55 27.8 21.4 32.0
61 23.5 22.8 34.2





#75
1,1,2,2-Tetrachloroethane
Concen: 20.902 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion: 83 Resp: 12069
Ion Ratio Lower Upper
83 100
131 8.3 5.0 15.0
85 63.7 31.9 95.7

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#76
1,2,3-Trichloropropane
Concen: 22.068 ug/l m
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion: 75 Resp: 111179
Ion Ratio Lower Upper
75 100
77 170.1 84.5 253.5





#77
Bromobenzene
Concen: 20.446 ug/l
RT: 12.988 min Scan# 1875
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion:156 Resp: 8152
Ion Ratio Lower Upper
156 100
77 232.0 107.7 323.3
158 95.3 48.4 145.3

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#78
n-propylbenzene
Concen: 19.667 ug/l
RT: 13.041 min Scan# 1885
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion: 91 Resp: 366798
Ion Ratio Lower Upper
91 100
120 22.1 11.5 34.4





#79
2-Chlorotoluene
Concen: 20.225 ug/l
RT: 13.129 min Scan# 1900
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion: 91 Resp: 240962
Ion Ratio Lower Upper
91 100
126 32.3 16.3 48.8

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#80
1,3,5-Trimethylbenzene
Concen: 20.286 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion:105 Resp: 257051
Ion Ratio Lower Upper
105 100
120 47.8 24.1 72.3





#81
trans-1,4-Dichloro-2-butene
Concen: 17.919 ug/l
RT: 12.741 min Scan# 1834
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion: 75 Resp: 33180
Ion Ratio Lower Upper
75 100
53 116.0 67.4 101.2
89 41.8 38.8 58.2

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#82
4-Chlorotoluene
Concen: 20.454 ug/l
RT: 13.229 min Scan# 1917
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion: 91 Resp: 242165
Ion Ratio Lower Upper
91 100
126 32.0 16.7 50.0





#83
tert-Butylbenzene
Concen: 20.178 ug/l
RT: 13.441 min Scan# 1953
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion:119 Resp: 209290
Ion Ratio Lower Upper
119 100
91 67.5 33.9 101.7
134 25.5 13.2 39.6

Manual Integrations
APPROVED


Reviewed By :John
Carlone

11/10/2023
Supervised By :Mahesh
Dadoda



11/10/2023




#84
1,2,4-Trimethylbenzene
Concen: 20.705 ug/l
RT: 13.488 min Scan# 1961
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59



Tgt Ion:105 Resp: 258119
Ion Ratio Lower Upper
105 100
120 44.5 22.0 66.0





#85
 sec-Butylbenzene
 Concen: 19.683 ug/l
 RT: 13.623 min Scan# 1984
 Delta R.T. 0.006 min
 Lab File: VN080026.D
 Acq: 09 Nov 2023 16:59

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1109WBSD01





Tgt Ion:105 Resp: 282849
 Ion Ratio Lower Upper
 105 100
 134 19.0 9.8 29.3

Manual Integrations
 APPROVED

Reviewed By :John
 Carlone




11/10/2023
 Supervised By :Mahesh
 Dadoda

11/10/2023



#86
 p-Isopropyltoluene
 Concen: 19.269 ug/l
 RT: 13.735 min Scan# 2003
 Delta R.T. 0.006 min
 Lab File: VN080026.D
 Acq: 09 Nov 2023 16:59


Tgt Ion:119 Resp: 225756
 Ion Ratio Lower Upper
 119 100
 134 25.8 12.6 37.8
 91 25.2 12.4 37.2





#87
1,3-Dichlorobenzene
Concen: 19.115 ug/l
RT: 13.735 min Scan# 2003
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion:146 Resp: 141788
Ion Ratio Lower Upper
146 100
111 42.9 22.1 66.3
148 64.5 32.0 96.0

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#88
1,4-Dichlorobenzene
Concen: 19.326 ug/l
RT: 13.818 min Scan# 2017
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion:146 Resp: 145505
Ion Ratio Lower Upper
146 100
111 42.4 20.6 61.8
148 64.2 31.6 94.8





#89
n-Butylbenzene
Concen: 18.825 ug/l
RT: 14.059 min Scan# 2058
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion: 91 Resp: 189958
Ion Ratio Lower Upper
91 100
92 51.0 26.4 79.0
134 22.8 12.6 37.6

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#90
Hexachloroethane
Concen: 19.604 ug/l
RT: 14.335 min Scan# 2105
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion:117 Resp: 45486
Ion Ratio Lower Upper
117 100
201 63.6 29.3 88.0





#91
1,2-Dichlorobenzene
Concen: 20.227 ug/l
RT: 14.112 min Scan# 2067
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion:146 Resp: 141629
Ion Ratio Lower Upper
146 100
111 45.2 22.0 66.0
148 63.8 31.8 95.3

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#92
1,2-Dibromo-3-Chloropropane
Concen: 21.580 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59


Tgt Ion: 75 Resp: 21132
Ion Ratio Lower Upper
75 100
155 67.4 39.1 117.2
157 89.8 50.3 150.9





#93
1,2,4-Trichlorobenzene
Concen: 18.749 ug/l
RT: 15.400 min Scan# 21
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01






Tgt Ion:180 Resp: 62424
Ion Ratio Lower Upper
180 100
182 92.7 46.1 138.2
145 30.9 16.6 49.8

Manual Integrations
APPROVED

Reviewed By :John
Carlone




11/10/2023
Supervised By :Mahesh
Dadoda


11/10/2023



#94
Hexachlorobutadiene
Concen: 19.839 ug/l
RT: 15.506 min Scan# 2304
Delta R.T. 0.000 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion:225 Resp: 29943
Ion Ratio Lower Upper
225 100
223 67.1 31.3 93.8
227 64.3 31.9 95.9





#95
Naphthalene
Concen: 15.955 ug/l
RT: 15.647 min Scan# 2327
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Instrument :
MSVOA_N
ClientSampleId :
VN1109WBSD01





Tgt Ion:128 Resp: 18014
Ion Ratio Lower Upper
128 100
127 13.1 10.0 15.0
129 11.4 8.7 13.1

Manual Integrations
APPROVED

Reviewed By :John
Carlone

11/10/2023
Supervised By :Mahesh
Dadoda

11/10/2023



#96
1,2,3-Trichlorobenzene
Concen: 18.829 ug/l
RT: 15.847 min Scan# 2362
Delta R.T. 0.006 min
Lab File: VN080026.D
Acq: 09 Nov 2023 16:59

Tgt Ion:180 Resp: 61451
Ion Ratio Lower Upper
180 100
182 94.5 46.2 138.6
145 34.2 17.5 52.6

