

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111020\
 Data File : VN064538.D
 Acq On : 10 Nov 2020 20:14
 Operator : JC/MD
 Sample : PB132908ZHE#04
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB132908ZHE#04

Quant Time: Nov 11 03:31:11 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	236983	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	488595	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	548847	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	186656	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	235815	57.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	114.26%	
35) Dibromofluoromethane	7.55	113	170663	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
50) Toluene-d8	10.06	98	668278	51.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.96%	
62) 4-Bromofluorobenzene	12.38	95	249661	51.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.68%	
Target Compounds						
16) Acetone	3.79	43	11071	7.041	ug/l	92
20) Methylene Chloride	4.51	84	8130	2.606	ug/l	88
43) Isopropyl Acetate	8.13	43	44585	5.330	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN111020\
Data File : VN064538.D
Acq On : 10 Nov 2020 20:14
Operator : JC/MD
Sample : PB132908ZHE#04
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB132908ZHE#04

Quant Time: Nov 11 03:31:11 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

