

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111020\
 Data File : VN064543.D
 Acq On : 10 Nov 2020 22:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 11 03:40:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	98	0.00
2 T	Dichlorodifluoromethane	0.651	0.630	3.2	90	0.00
3 P	Chloromethane	0.791	0.762	3.7	92	0.00
4 C	Vinyl Chloride	0.812	0.830	-2.2#	96	0.00
5 T	Bromomethane	0.519	0.519	0.0	93	0.00
6 T	Chloroethane	0.520	0.521	-0.2	95	0.00
7 T	Trichlorofluoromethane	1.195	1.196	-0.1	93	0.00
8 T	Diethyl Ether	0.422	0.425	-0.7	95	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.568	0.562	1.1	94	0.00
10 T	Methyl Iodide	0.660	0.664	-0.6	91	0.00
11 T	Tert butyl alcohol	0.096	0.095	1.0	89	0.00
12 CM	1,1-Dichloroethene	0.535	0.534	0.2#	93	0.00
13 T	Acrolein	0.060	0.089	-48.3#	146	0.00
14 T	Allyl chloride	1.046	1.011	3.3	89	0.00
15 T	Acrylonitrile	0.267	0.292	-9.4	94	0.00
16 T	Acetone	0.332	0.268	19.3	74	0.00
17 T	Carbon Disulfide	1.599	1.553	2.9	90	0.00
18 T	Methyl Acetate	0.652	0.754	-15.6	112	0.00
19 T	Methyl tert-butyl Ether	1.925	1.939	-0.7	91	0.00
20 T	Methylene Chloride	0.658	0.651	1.1	96	0.00
21 T	trans-1,2-Dichloroethene	0.607	0.606	0.2	92	0.00
22 T	Diisopropyl ether	2.220	2.259	-1.8	91	0.00
23 T	Vinyl Acetate	1.806	1.855	-2.7	89	0.00
24 P	1,1-Dichloroethane	1.220	1.224	-0.3	92	0.00
25 T	2-Butanone	0.423	0.412	2.6	85	0.00
26 T	2,2-Dichloropropane	1.105	0.934	15.5	78	0.00
27 T	cis-1,2-Dichloroethene	0.704	0.719	-2.1	94	0.00
28 T	Bromochloromethane	0.600	0.585	2.5	108	0.00
29 T	Tetrahydrofuran	0.259	0.273	-5.4	94	0.00
30 C	Chloroform	1.283	1.307	-1.9#	93	0.00
31 T	Cyclohexane	1.059	0.975	7.9	88	0.00
32 T	1,1,1-Trichloroethane	1.138	1.154	-1.4	94	0.00
33 S	1,2-Dichloroethane-d4	0.871	0.819	6.0	96	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.340	0.328	3.5	100	0.00
36 T	1,1-Dichloropropene	0.528	0.518	1.9	90	0.00
37 T	Ethyl Acetate	0.472	0.503	-6.6	91	0.00
38 T	Carbon Tetrachloride	0.565	0.566	-0.2	93	0.00
39 T	Methylcyclohexane	0.474	0.463	2.3	89	0.00
40 TM	Benzene	1.552	1.551	0.1	94	0.00
41 T	Methacrylonitrile	0.234	0.227	3.0	100	0.00
42 TM	1,2-Dichloroethane	0.626	0.630	-0.6	93	0.00
43 T	Isopropyl Acetate	0.856	0.864	-0.9	92	0.00
44 TM	Trichloroethene	0.392	0.388	1.0	93	0.00
45 C	1,2-Dichloropropane	0.414	0.421	-1.7#	94	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111020\
 Data File : VN064543.D
 Acq On : 10 Nov 2020 22:14
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 11 03:40:32 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 04 12:09:03 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.273	0.277	-1.5	95	0.00
47 T	Bromodichloromethane	0.586	0.597	-1.9	95	0.00
48 T	Methyl methacrylate	0.432	0.449	-3.9	94	0.00
49 T	1,4-Dioxane	0.006	0.006	0.0	95	0.00
50 S	Toluene-d8	1.316	1.229	6.6	96	0.00
51 T	4-Methyl-2-Pentanone	0.499	0.528	-5.8	93	0.00
52 CM	Toluene	0.947	0.954	-0.7#	92	0.00
53 T	t-1,3-Dichloropropene	0.621	0.622	-0.2	88	0.00
54 T	cis-1,3-Dichloropropene	0.664	0.665	-0.2	90	0.00
55 T	1,1,2-Trichloroethane	0.369	0.377	-2.2	94	0.00
56 T	Ethyl methacrylate	0.535	0.566	-5.8	93	0.00
57 T	1,3-Dichloropropane	0.644	0.667	-3.6	95	0.00
58 T	2-Chloroethyl Vinyl ether	0.253	0.292	-15.4	129	0.00
59 T	2-Hexanone	0.362	0.380	-5.0	90	0.00
60 T	Dibromochloromethane	0.413	0.440	-6.5	95	0.00
61 T	1,2-Dibromoethane	0.374	0.393	-5.1	94	0.00
62 S	4-Bromofluorobenzene	0.498	0.461	7.4	93	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	100	0.00
64 T	Tetrachloroethene	0.385	0.364	5.5	89	0.00
65 PM	Chlorobenzene	1.096	1.073	2.1	92	0.00
66 T	1,1,1,2-Tetrachloroethane	0.409	0.423	-3.4	96	0.00
67 C	Ethyl Benzene	1.950	1.977	-1.4#	91	0.00
68 T	m/p-Xylenes	0.707	0.733	-3.7	92	0.00
69 T	o-Xylene	0.683	0.696	-1.9	91	0.00
70 T	Styrene	1.150	1.227	-6.7	93	0.00
71 P	Bromoform	0.276	0.307	-11.2	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	102	0.00
73 T	Isopropylbenzene	3.974	3.897	1.9	93	0.00
74 T	N-amyl acetate	1.727	1.746	-1.1	91	0.00
75 P	1,1,2,2-Tetrachloroethane	1.221	1.222	-0.1	100	0.00
76 T	1,2,3-Trichloropropane	1.211	1.153	4.8	90	0.00
77 T	Bromobenzene	0.970	0.931	4.0	95	0.00
78 T	n-propylbenzene	4.531	4.500	0.7	93	0.00
79 T	2-Chlorotoluene	2.874	2.798	2.6	95	0.00
80 T	1,3,5-Trimethylbenzene	3.279	3.266	0.4	93	0.00
81 T	trans-1,4-Dichloro-2-butene	0.434	0.400	7.8	86	0.00
82 T	4-Chlorotoluene	3.009	2.924	2.8	92	0.00
83 T	tert-Butylbenzene	2.561	2.588	-1.1	96	0.00
84 T	1,2,4-Trimethylbenzene	3.296	3.353	-1.7	93	0.00
85 T	sec-Butylbenzene	3.243	3.328	-2.6	94	0.00
86 T	p-Isopropyltoluene	2.917	3.052	-4.6	95	0.00
87 T	1,3-Dichlorobenzene	1.677	1.654	1.4	94	0.00
88 T	1,4-Dichlorobenzene	1.733	1.666	3.9	95	0.00
89 T	n-Butylbenzene	2.508	2.550	-1.7	92	0.00

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN111020\
Data File : VN064543.D
Acq On : 10 Nov 2020 22:14
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Nov 11 03:40:32 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N110420W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 04 12:09:03 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.563	0.587	-4.3	98	0.00
91 T	1,2-Dichlorobenzene	1.613	1.686	-4.5	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.243	0.256	-5.3	97	0.00
93 T	1,2,4-Trichlorobenzene	0.660	0.735	-11.4	94	0.00
94 T	Hexachlorobutadiene	0.328	0.348	-6.1	102	0.00
95 T	Naphthalene	2.152	2.410	-12.0	94	0.00
96 T	1,2,3-Trichlorobenzene	0.654	0.738	-12.8	100	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6