

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111021\
 Data File : VN069443.D
 Acq On : 10 Nov 2021 15:41
 Operator : JC/MD
 Sample : VN1110WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
ClientSampleId :
 VN1110WBL01

Quant Time: Nov 10 18:04:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N111021W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Nov 10 14:36:17 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Bromochloromethane	7.662	128	162804	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	8.971	114	962420	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.747	117	814145	30.000	ug/l	0.00
System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.440	65	453030	32.085	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	106.967%
60) 4-Bromofluorobenzene	12.734	95	316738	24.311	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	81.033%
63) Toluene-d8	10.443	98	1178839	31.359	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	104.533%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111021\
 Data File : VN069443.D
 Acq On : 10 Nov 2021 15:41
 Operator : JC/MD
 Sample : VN1110WBL01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN1110WBL01

Quant Time: Nov 10 18:04:18 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N111021W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Wed Nov 10 14:36:17 2021
 Response via : Initial Calibration

