

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111022\
 Data File : VN075232.D
 Acq On : 10 Nov 2022 14:10
 Operator : JC\MD
 Sample : PB148933TB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB148933TB

Quant Time: Nov 11 04:10:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	106183	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	197430	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	183002	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	76936	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	74972	52.403	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	104.800%
35) Dibromofluoromethane	8.171	113	63755	55.008	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	110.020%
50) Toluene-d8	10.571	98	232244	51.648	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	103.300%
62) 4-Bromofluorobenzene	12.847	95	80906	51.220	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	102.440%
Target Compounds						
					Qvalue	
16) Acetone	4.453	43	9545	15.756	ug/l	97
20) Methylene Chloride	5.289	84	5781	4.280	ug/l #	87
37) Ethyl Acetate	7.565	43	17449	8.867	ug/l	100
43) Isopropyl Acetate	8.694	43	77522	25.397	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111022\
Data File : VN075232.D
Acq On : 10 Nov 2022 14:10
Operator : JC\MD
Sample : PB148933TB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB148933TB

Quant Time: Nov 11 04:10:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

