

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111022\
 Data File : VN075233.D
 Acq On : 10 Nov 2022 14:34
 Operator : JC\MD
 Sample : N5537-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 360

Quant Time: Nov 11 04:10:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 27 09:48:49 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	101625	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	194542	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	180626	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	76012	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	75769	55.335	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 110.680%	
35) Dibromofluoromethane	8.171	113	62157	54.425	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 108.860%	
50) Toluene-d8	10.565	98	228919	51.664	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 103.320%	
62) 4-Bromofluorobenzene	12.847	95	80276	51.575	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 103.160%	
Target Compounds						
					Qvalue	
16) Acetone	4.448	43	13380	23.077	ug/l	96
18) Methyl Acetate	5.048	43	2482	1.311	ug/l #	83
20) Methylene Chloride	5.271	84	8620	6.668	ug/l	85
37) Ethyl Acetate	7.571	43	16452	8.485	ug/l	97
43) Isopropyl Acetate	8.694	43	73589	24.467	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111022\
Data File : VN075233.D
Acq On : 10 Nov 2022 14:34
Operator : JC\MD
Sample : N5537-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
360

Quant Time: Nov 11 04:10:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N102622W.M
Quant Title : SW846 8260
QLast Update : Thu Oct 27 09:48:49 2022
Response via : Initial Calibration

