

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111023\
 Data File : VN080047.D
 Acq On : 10 Nov 2023 12:48
 Operator : JC\MD
 Sample : 05325-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 231108054-03-TRIP-BLANK

Quant Time: Nov 10 23:56:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N110823W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Nov 09 04:02:05 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.818	128	78109	30.000	ug/l	0.00
28) 1,4-Difluorobenzene	9.106	114	394070	30.000	ug/l	0.00
57) Chlorobenzene-d5	11.871	117	356439	30.000	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	8.583	65	200327	30.553	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	101.833%
60) 4-Bromofluorobenzene	12.853	95	139614	24.494	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	81.633%
63) Toluene-d8	10.571	98	488056	28.898	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	96.333%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN111023\
Data File : VN080047.D
Acq On : 10 Nov 2023 12:48
Operator : JC\MD
Sample : 05325-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
231108054-03-TRIP-BLANK

Quant Time: Nov 10 23:56:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N110823W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Nov 09 04:02:05 2023
Response via : Initial Calibration









#1
Bromochloromethane
Concen: 30.000 ug/l
RT: 7.818 min Scan# 91
Delta R.T. -0.006 min
Lab File: VN080047.D
Acq: 10 Nov 2023 12:48




Instrument :
MSVOA_N
ClientSampleId :
231108054-03-TRIP-BLANK


Tgt Ion:128 Resp: 78109
Ion Ratio Lower Upper
128 100
49 182.8 0.0 468.5
130 133.3 0.0 320.7



#27
1,2-Dichloroethane-d4
Concen: 30.553 ug/l
RT: 8.583 min Scan# 1127
Delta R.T. 0.001 min
Lab File: VN080047.D
Acq: 10 Nov 2023 12:48

Tgt Ion: 65 Resp: 200327
Ion Ratio Lower Upper
65 100
67 52.5 42.1 63.1





#28

1,4-Difluorobenzene

Concen: 30.000 ug/l

RT: 9.106 min Scan# 1

Delta R.T. 0.000 min

Lab File: VN080047.D

Acq: 10 Nov 2023 12:48

Instrument :

MSVOA_N

ClientSampleId :

231108054-03-TRIP-BLANK





Tgt Ion:114 Resp: 394070

Ion Ratio Lower Upper

114 100

63 21.7 18.1 27.1

88 16.3 12.2 18.2



#57

Chlorobenzene-d5

Concen: 30.000 ug/l

RT: 11.871 min Scan# 1686

Delta R.T. 0.001 min

Lab File: VN080047.D

Acq: 10 Nov 2023 12:48




Tgt Ion:117 Resp: 356439


Ion Ratio Lower Upper

117 100

82 58.3 46.2 69.2

119 32.4 25.2 37.8









#60
4-Bromofluorobenzene
Concen: 24.494 ug/l
RT: 12.853 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN080047.D
Acq: 10 Nov 2023 12:48

Instrument :
MSVOA_N
ClientSampleId :
231108054-03-TRIP-BLANK

Tgt Ion: 95 Resp: 139614
Ion Ratio Lower Upper
95 100
174 72.3 60.5 90.7
176 69.2 57.2 85.8



#63
Toluene-d8
Concen: 28.898 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. 0.001 min
Lab File: VN080047.D
Acq: 10 Nov 2023 12:48

Tgt Ion: 98 Resp: 488056
Ion Ratio Lower Upper
98 100
100 65.3 51.6 77.4

